



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110175

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	3.99	07/23/2020	17:30
2	CAL2	1	Water	NA	NA	20.3	7.01	07/23/2020	17:31
3	CAL3	1	Water	NA	NA	20.3	10.03	07/23/2020	17:32
4	ICV	1	Water	NA	NA	20.3	6.99	07/23/2020	17:35
5	CCV1	1	Water	NA	NA	20.3	2.01	07/23/2020	17:36
6	L3411-01	1	Solid	20.03	20	25.6	7.77	07/23/2020	17:28
7	L3411-01DUP	1	Solid	20.04	20	25.7	7.78	07/23/2020	17:29
8	L3411-03	1	Solid	20.03	20	25.6	7.26	07/23/2020	17:30
9	L3411-04	1	Solid	20.02	20	25.5	6.80	07/23/2020	17:31
10	CCV2	1	Water	NA	NA	20.3	12.02	07/23/2020	17:33

✓B110175

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13411 WorkList ID : 140757 Department : Wet-Chemistry Date : 07-23-2020 16:58:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2020	Solid	L3411-01	pH	Cool 4 deg C	PSEG01	H12	COMP-31076-31074	07/23/2020	9045D
07/23/2020	Solid	L3411-03	pH	Cool 4 deg C	PSEG01	H12	31076	07/23/2020	9045D
07/23/2020	Solid	L3411-04	pH	Cool 4 deg C	PSEG01	H12	31074	07/23/2020	9045D

Date/Time 07/23/2020 17:00
Received by: JP W11-0161M
Relinquished by: KS (wetchem)

Date/Time 07/23/2020 18:10
Received by: KS (wetchem)
Relinquished by: JP W11-0161M