



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110176

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	07/23/2020	17:30
2	CAL2	1	Water	NA	NA	20.2	7.01	07/23/2020	17:31
3	CAL3	1	Water	NA	NA	20.3	10.03	07/23/2020	17:32
4	ICV	1	Water	NA	NA	20.3	7.00	07/23/2020	17:35
5	CCV1	1	Water	NA	NA	20.3	2.01	07/23/2020	17:36
6	L3410-04	1	Water	NA	NA	22.2	7.14	07/23/2020	17:38
7	L3410-04DUP	1	Water	NA	NA	22.2	7.15	07/23/2020	17:39
8	CCV2	1	Water	NA	NA	20.3	12.02	07/23/2020	17:40

28110176

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3410 WorkList ID : 140754 Department : Wet-Chemistry Date : 07-23-2020 16:56:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2020	Water	L3410-04	Corrosivity	Cool 4 deg C	PSEG01	I13	FM-FRAC	07/23/2020	9040C

Date/Time 07/23/2020 17:00
Received by: JP (wet-chem)
Relinquished by: KS (wet-chem)

Date/Time 07/23/2020 18:10
Received by: KS (wet-chem)
Relinquished by: JP (wet-chem)