



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110181

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/23/2020	17:30
2	CAL2	1	Water	NA	NA	20.3	7.00	07/23/2020	17:31
3	CAL3	1	Water	NA	NA	20.4	10.03	07/23/2020	17:32
4	ICV	1	Water	NA	NA	20.4	7.00	07/23/2020	17:35
5	CCV1	1	Water	NA	NA	20.4	2.02	07/23/2020	17:36
6	L3410-03	1	Water	NA	NA	22.1	7.14	07/23/2020	17:37
7	L3410-03DUP	1	Water	NA	NA	22.2	7.15	07/23/2020	17:38
8	L3411-05	1	Water	NA	NA	23.5	7.76	07/23/2020	17:40
9	L3413-07	1	Water	NA	NA	25.3	5.06	07/23/2020	17:42
10	L3413-08	1	Water	NA	NA	25.2	4.72	07/23/2020	17:45
11	CCV2	1	Water	NA	NA	20.4	12.02	07/23/2020	17:46

VB 110181

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13410 WorkList ID : 140756 Department : Wet-Chemistry Date : 07-23-2020 16:57:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2020	Water	L3410-03	pH	Cool 4 deg C	PSEG01	I13	FM-FRAC	07/23/2020	9040C
07/23/2020	Water	L3411-05	pH	Cool 4 deg C	PSEG01	H12	OILY-WASH-WATER	07/23/2020	9040C
07/23/2020	Water	L3413-07	pH	Cool 4 deg C	PSEG01	I11	20758	07/23/2020	9040C
07/23/2020	Water	L3413-08	pH	Cool 4 deg C	PSEG01	I11	40956	07/23/2020	9040C

Date/Time 07/23/2020 17:00
 Received by: JB wyl-ehens
 Relinquished by: JB

Date/Time 07/23/2020 18:10
 Received by: JB
 Relinquished by: JB wyl-ehens