

LB110183


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46265



A= 0.0000e+000

B= 2.1319e-004

C= 1.6896e-002

Rho= 0.9994570

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	93	0.000	93				
0.2	0.200	0.241	0.041	1049	0.0 %	1049				
2.5	2.500	2.301	-0.199	10714	0.0 %	10714				
5.0	5.000	5.085	0.085	23773	0.0 %	23773				
7.5	7.500	7.657	0.157	35835	0.0 %	35835				
10.0	10.000	9.880	-0.120	46265	0.0 %	46265				

LB110183 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jul 2020 11:33:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	93	-	0.0000	7471B	PPB	24 Jul 2020 11:35:03	S	Std
0.2 - 1	50.2	1049	-	0.2000	7471B	PPB	24 Jul 2020 11:37:20	S	Std
2.5 - 1	52.5	10714	-	2.5000	7471B	PPB	24 Jul 2020 11:39:37	S	Std
5.0 - 1	55	23773	-	5.0000	7471B	PPB	24 Jul 2020 11:44:29	S	Std
7.5 - 1	57.5	35835	-	7.5000	7471B	PPB	24 Jul 2020 11:49:20	S	Std
10.0 - 1	518	46265	-	10.0000	7471B	PPB	24 Jul 2020 11:51:38	S	Std
ICV40 - 1	ICV40	18443	3.9488	-	7471B	PPB	24 Jul 2020 11:54:31	U	SMPL
ICB40 - 1	ICB40	-279	-0.0426	-	7471B	PPB	24 Jul 2020 11:56:46	U	SMPL
CCV60 - 1	CCV60	23687	5.0667	-	7471B	PPB	24 Jul 2020 11:59:03	U	SMPL
CCB60 - 1	CCB60	-276	-0.0419	-	7471B	PPB	24 Jul 2020 12:01:18	U	SMPL
CRA - 1	CRA	829	0.1936	-	7471B	PPB	24 Jul 2020 12:03:35	U	SMPL
HighStd - 1	HighStd	46118	9.8488	-	7471B	PPB	24 Jul 2020 12:05:50	U	SMPL
ChkStd - 1	ChkStd	34957	7.4694	-	7471B	PPB	24 Jul 2020 12:08:05	U	SMPL
PB130440BL - 1	PBS	-241	-0.0345	-	7471B	PPB	24 Jul 2020 12:10:24	U	SMPL
PB130440BS - 1	LCSS	19780	4.2338	-	7471B	PPB	24 Jul 2020 12:12:43	U	SMPL
L3382-01 - 1	CTR-007	5511	1.1918	-	7471B	PPB	24 Jul 2020 12:15:00	U	SMPL
L3382-01DUP - 1	CTR-007DUP	4928	1.0675	-	7471B	PPB	24 Jul 2020 12:17:18	U	SMPL
L3382-01MS - 1	CTR-007MS	22449	4.8028	-	7471B	PPB	24 Jul 2020 12:19:38	U	SMPL
L3382-01MSD - 1	CTR-007MSD	23867	5.1051	-	7471B	PPB	24 Jul 2020 12:21:57	U	SMPL
L3382-05 - 1	CTR-020	165	0.0521	-	7471B	PPB	24 Jul 2020 12:24:15	U	SMPL
CCV61 - 1	CCV61	23027	4.9260	-	7471B	PPB	24 Jul 2020 12:26:35	U	SMPL
CCB61 - 1	CCB61	-324	-0.0522	-	7471B	PPB	24 Jul 2020 12:28:52	U	SMPL
L3382-09 - 1	CTR-023	4258	0.9247	-	7471B	PPB	24 Jul 2020 12:31:09	U	SMPL
L3382-13 - 1	CTR-005	4671	1.0127	-	7471B	PPB	24 Jul 2020 12:33:24	U	SMPL
L3382-17 - 1	CTR-011	6316	1.3634	-	7471B	PPB	24 Jul 2020 12:35:41	U	SMPL
L3382-21 - 1	CTR-024	1558	0.3490	-	7471B	PPB	24 Jul 2020 12:37:59	U	SMPL
L3383-01 - 1	CRT-013	4294	0.9323	-	7471B	PPB	24 Jul 2020 12:40:16	U	SMPL
L3383-05 - 1	CRT-018	2617	0.5748	-	7471B	PPB	24 Jul 2020 12:42:34	U	SMPL
L3383-09 - 1	CRT-009	4272	0.9276	-	7471B	PPB	24 Jul 2020 12:44:53	U	SMPL
L3383-13 - 1	CRT-022	10114	2.1731	-	7471B	PPB	24 Jul 2020 12:47:11	U	SMPL
L3383-17 - 1	CRT-008	1248	0.2830	-	7471B	PPB	24 Jul 2020 12:49:29	U	SMPL
L3383-21 - 1	CRT-015	1400	0.3154	-	7471B	PPB	24 Jul 2020 12:51:46	U	SMPL
CCV62 - 1	CCV62	23297	4.9836	-	7471B	PPB	24 Jul 2020 12:54:02	U	SMPL
CCB62 - 1	CCB62	-263	-0.0392	-	7471B	PPB	24 Jul 2020 12:56:17	U	SMPL
L3383-25 - 1	CRT-014	2372	0.5226	-	7471B	PPB	24 Jul 2020 12:58:35	U	SMPL
L3383-29 - 1	CRT-002	5923	1.2796	-	7471B	PPB	24 Jul 2020 13:00:50	U	SMPL
L3408-02 - 1	PS-DRUM	20786	4.4483	-	7471B	PPB	24 Jul 2020 13:03:07	U	SMPL
L3409-02 - 1	354	13995	3.0005	-	7471B	PPB	24 Jul 2020 13:05:23	U	SMPL
L3410-01 - 1	FM-CON-A	20927	4.4783	-	7471B	PPB	24 Jul 2020 13:07:41	U	SMPL
L3410-02 - 1	FM-CON-B	27308	5.8387	-	7471B	PPB	24 Jul 2020 13:09:59	U	SMPL
L3412-01 - 1	DUCK-BANK	4803	1.0408	-	7471B	PPB	24 Jul 2020 13:12:18	U	SMPL
L3414-01 - 1	61820	3372	0.7358	-	7471B	PPB	24 Jul 2020 13:14:37	U	SMPL
L3382-01LX5 - 1		781	0.1834	-	7471B	PPB	24 Jul 2020 13:16:56	U	SMPL
L3382-01A - 1		23024	4.9254	-	7471B	PPB	24 Jul 2020 13:19:12	U	SMPL
CCV63 - 1	CCV63	22851	4.8885	-	7471B	PPB	24 Jul 2020 13:21:27	U	SMPL
CCB63 - 1	CCB63	-259	-0.0383	-	7471B	PPB	24 Jul 2020 13:23:45	U	SMPL