



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: Alexander

Parameter: Color

Supervisor Review By: apatel

Run Number: LB110185

pH Meter ID: WC PH METER-1

Reagent/Standard	Lot/Log #
30 color units	WP85142
25 color units	WP85141
20 color units	WP85140
15 color units	WP85139
5 color units	WP85138
70 color units	WP85146
60 color units	WP85145
50 color units	WP85144
40 color units	WP85143
0 color units	WP85136
10 color units	WP85137

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB110185BL	LB110185BL	1	5.19	N/A	0.00	07/24/2020	09:20
2	L3421-02	FIELD-BLANK	1	4.57	N/A	0.00	07/24/2020	09:23
3	L3421-03	SY-6	5	6.46	N/A	50.00	07/24/2020	09:26
4	L3421-03DUP	SY-6DUP	5	6.49	N/A	50.00	07/24/2020	09:29
5	L3421-04	SY-3	10	6.14	N/A	60.00	07/24/2020	09:32
6	L3421-05	SY-3D	10	6.48	N/A	60.00	07/24/2020	09:35
7	L3421-06	SY-3DD	1	6.14	N/A	0.00	07/24/2020	09:38
8	L3421-07	SY-7	1	6.04	N/A	0.00	07/24/2020	09:40

WORKLIST(Hardcopy Internal Chain)

LB110185

WorkList Name : Color-72420

WorkList ID : 140768

Department : Wet-Chemistry

Date : 07-24-2020 08:55:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2020	Water	L3421-02 <i>I</i>	Color	Cool 4 deg C	LOCK01	I12	FIELD-BLANK	07/22/2020	SM2120 B
07/24/2020	Water	L3421-03 <i>D</i>	Color	Cool 4 deg C	LOCK01	I12	SY-6	07/22/2020	SM2120 B
07/24/2020	Water	L3421-04 <i>D</i>	Color	Cool 4 deg C	LOCK01	I12	SY-3	07/22/2020	SM2120 B
07/24/2020	Water	L3421-05 <i>I</i>	Color	Cool 4 deg C	LOCK01	I12	SY-3D	07/22/2020	SM2120 B
07/24/2020	Water	L3421-06 <i>I</i>	Color	Cool 4 deg C	LOCK01	I12	SY-3DD	07/22/2020	SM2120 B
07/24/2020	Water	L3421-07 <i>I</i>	Color	Cool 4 deg C	LOCK01	I12	SY-7	07/22/2020	SM2120 B

Date/Time 7/24/2020 9:05
 Received by: AM (wetchem)
 Relinquished by: JA (wetchem)

Date/Time 7/24/2020 9:50
 Received by: JA (wetchem)
 Relinquished by: AM (wetchem)