



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110186

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/24/2020	10:05
2	L3358-01		1	8	100.00	>212.0	>212.0	07/24/2020	10:35
3	L3362-05		1	9	100.00	>212.0	>212.0	07/24/2020	11:05
4	L3363-03		1	9	100.00	>212.0	>212.0	07/24/2020	11:40
5	L3363-05		1	8	72.00	161.6	161.6	07/24/2020	12:10
6	L3363-05DUP		1	8	72.00	161.6	161.6	07/24/2020	12:40
7	L3370-10		1	9	100.00	>212.0	>212.0	07/24/2020	13:05
8	L3378-17		1	8	100.00	>212.0	>212.0	07/24/2020	13:32
9	L3410-04		1	7	100.00	>212.0	>212.0	07/24/2020	13:58
10	L3411-05		1	8	100.00	>212.0	>212.0	07/24/2020	14:20
11	L3413-05		1	8	100.00	>212.0	>212.0	07/24/2020	14:45
12	L3413-06		1	9	100.00	>212.0	>212.0	07/24/2020	15:10
13	L3413-07		1	8	100.00	>212.0	>212.0	07/24/2020	15:35
14	L3413-08		1	7	100.00	>212.0	>212.0	07/24/2020	16:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp72220 WorkList ID : 140673 Department : Wet-Chemistry Date : 07-22-2020 09:29:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/03/2020	Water	L3358-01 G	Flash Point	Cool 4 deg C	DIST01		55-ST-28	07/20/2020	1010A
07/28/2020	Water	L3362-05 H	Flash Point	Cool 4 deg C	PSEG01	H31	1754	07/21/2020	1010A
07/28/2020	Water	L3363-03 A	Flash Point	Cool 4 deg C	PSEG01	H31	79	07/21/2020	1010A
07/28/2020	Water	L3363-05 B	Flash Point	Cool 4 deg C	PSEG01	H31	80	07/21/2020	1010A
07/28/2020	Water	L3370-10 C	Flash Point	Cool 4 deg C	PSEG01	H31	TP-3-GW	07/21/2020	1010A
07/28/2020	Water	L3378-17 A	Flash Point	Cool 4 deg C	PSEG01	H43	DRUM-WATER	07/22/2020	1010A
07/28/2020	Water	L3410-04 C	Flash Point	Cool 4 deg C	PSEG01	I13	FM-FRAC	07/23/2020	1010A
08/06/2020	Water	L3411-05 A	Flash Point	Cool 4 deg C	PSEG01	H12	OILY-WASH-WATER	07/23/2020	1010A
07/28/2020	Water	L3413-05 B	Flash Point	Cool 4 deg C	PSEG01	I11	40394	07/23/2020	1010A
07/30/2020	Water	L3413-06 B	Flash Point	Cool 4 deg C	PSEG01	I11	40888	07/23/2020	1010A
07/30/2020	Water	L3413-07 B	Flash Point	Cool 4 deg C	PSEG01	I11	20758	07/23/2020	1010A
07/30/2020	Water	L3413-08 B	Flash Point	Cool 4 deg C	PSEG01	I11	40956	07/23/2020	1010A

Date/Time 07/24/2020 09:35
 Received by: KS (wet-chem)
 Relinquished by: JP (wet-chem)

Date/Time 07/24/2020 17:00
 Received by: JP (wet-chem)
 Relinquished by: KS (wet-chem)