


LB110198

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5734e-004

C= -1.4138e-002

Rho= 0.9999818

Accept=Accepted

Slope
Y=101600x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.011	0.011	97	0.000	97					-2
0.05	0.050	0.049	-0.001	247	0.0 %	247					-12
0.20	0.200	0.177	-0.023	743	0.0 %	743					2
2.50	2.500	2.539	0.039	9920	0.0 %	9920					0
5.00	5.000	4.979	-0.021	19403	0.0 %	19403					0
7.50	7.500	7.476	-0.024	29105	0.0 %	29105					0
10.0	10.000	10.019	0.019	38987	0.0 %	38987					0

LB110198
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.05	97	-		0 ISM02.4	PPB	7/24/2020 16:49	Std
	0.05 S0.05	247	-	0.05	ISM02.4	PPB	7/24/2020 16:51	Std
	0.2 S0.2	743	-	0.2	ISM02.4	PPB	7/24/2020 16:56	Std
	2.5 S2.5	9920	-	2.5	ISM02.4	PPB	7/24/2020 16:59	Std
	5 S5.0	19403	-	5	ISM02.4	PPB	7/24/2020 17:02	Std
	7.5 S7.5	29105	-	7.5	ISM02.4	PPB	7/24/2020 17:05	Std
	10 S10.0	38987	-	10	ISM02.4	PPB	7/24/2020 17:18	Std
ICV	ICV	14711	3.7717 -		ISM02.4	PPB	7/24/2020 17:20	SMPL
ICB	ICB	-23	-0.0201 -		ISM02.4	PPB	7/24/2020 17:25	SMPL
CCV067	CCV067	19680	5.0504 -		ISM02.4	PPB	7/24/2020 17:27	SMPL
CCB067	CCB067	29	-0.0067 -		ISM02.4	PPB	7/24/2020 17:32	SMPL
PB130476BL	PBW476	10	-0.0116 -		ISM02.4	PPB	7/24/2020 17:35	SMPL
L3384-01	MBFS44	29356	7.5405 -		ISM02.4	PPB	7/24/2020 17:37	SMPL
L3384-02	MBFS46	2163	0.5425 -		ISM02.4	PPB	7/24/2020 17:39	SMPL
L3384-03	MBFS48	1252	0.3081 -		ISM02.4	PPB	7/24/2020 17:54	SMPL
L3384-04	MBFS48D	1044	0.2545 -		ISM02.4	PPB	7/24/2020 17:56	SMPL
L3384-05	MBFS48S	5074	1.2916 -		ISM02.4	PPB	7/24/2020 17:58	SMPL
CCV068	CCV068	19102	4.9017 -		ISM02.4	PPB	7/24/2020 18:01	SMPL
CCB068	CCB068	5	-0.0129 -		ISM02.4	PPB	7/24/2020 18:03	SMPL
PB130477BL	PBW477	8	-0.0121 -		ISM02.4	PPB	7/24/2020 18:05	SMPL
L3385-01	MBFS45	112	0.0147 -		ISM02.4	PPB	7/24/2020 18:08	SMPL
L3385-02	MBFS47	14	-0.0105 -		ISM02.4	PPB	7/24/2020 18:10	SMPL
L3385-03	MBFS49	540	0.1248 -		ISM02.4	PPB	7/24/2020 18:17	SMPL
L3385-04	MBFS49D	452	0.1022 -		ISM02.4	PPB	7/24/2020 18:19	SMPL
L3385-05	MBFS49S	4479	1.1385 -		ISM02.4	PPB	7/24/2020 18:21	SMPL
CCV069	CCV069	18905	4.851 -		ISM02.4	PPB	7/24/2020 18:24	SMPL
CCB069	CCB069	64	0.0023 -		ISM02.4	PPB	7/24/2020 18:26	SMPL