



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110200

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/24/2020	17:10
2	CAL2	1	Water	NA	NA	20.1	7.01	07/24/2020	17:11
3	CAL3	1	Water	NA	NA	20.3	10.03	07/24/2020	17:12
4	ICV	1	Water	NA	NA	21.4	7.00	07/24/2020	17:15
5	CCV1	1	Water	NA	NA	20.3	2.01	07/24/2020	17:16
6	L3428-04	1	Solid	20.03	20	24.8	7.25	07/24/2020	17:20
7	L3428-04DUP	1	Solid	20.04	20	24.9	7.26	07/24/2020	17:21
8	L3434-02	1	Solid	20.03	20	24.8	6.43	07/24/2020	17:22
9	L3434-04	1	Solid	20.04	20	24.8	7.85	07/24/2020	17:25
10	L3439-08	1	Solid	20.02	20	24.2	11.20	07/24/2020	17:26
11	L3439-09	1	Solid	20.03	20	24.1	11.16	07/24/2020	17:29
12	L3439-10	1	Solid	20.05	20	24.0	11.27	07/24/2020	17:30
13	L3439-11	1	Solid	20.02	20	23.8	11.19	07/24/2020	17:32
14	L3439-12	1	Solid	20.03	20	23.8	11.10	07/24/2020	17:35
15	L3439-13	1	Solid	20.04	20	23.6	11.36	07/24/2020	17:36
16	CCV2	1	Water	NA	NA	20.4	12.02	07/24/2020	17:38
17	L3439-14	1	Solid	20.02	20	23.7	11.17	07/24/2020	17:40
18	CCV3	1	Water	NA	NA	20.3	2.02	07/24/2020	17:42

116200

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3479

WorkList ID : 140804

Department : Wet-Chemistry

Date : 07-24-2020 13:45:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2020	Solid	L3428-04	Corrosivity	Cool 4 deg C	PSEG01	I12	TP-1-S	07/23/2020	9045D
07/24/2020	Solid	L3434-02	Corrosivity	Cool 4 deg C	PSEG01	I23	11986	07/24/2020	9045D
07/24/2020	Solid	L3434-04	Corrosivity	Cool 4 deg C	PSEG01	I23	11936	07/24/2020	9045D
07/23/2020	Solid	L3439-08	Corrosivity	Cool 4 deg C	AQUA04	I13	20200438-3001A	07/23/2020	9045D
07/23/2020	Solid	L3439-09	Corrosivity	Cool 4 deg C	AQUA04	I13	20200439-3001B	07/23/2020	9045D
07/23/2020	Solid	L3439-10	Corrosivity	Cool 4 deg C	AQUA04	I13	20200440-3001C	07/23/2020	9045D
07/23/2020	Solid	L3439-11	Corrosivity	Cool 4 deg C	AQUA04	I13	20200441-3002A	07/23/2020	9045D
07/23/2020	Solid	L3439-12	Corrosivity	Cool 4 deg C	AQUA04	I13	20200442-3002B	07/23/2020	9045D
07/23/2020	Solid	L3439-13	Corrosivity	Cool 4 deg C	AQUA04	I13	20200443-3002C	07/23/2020	9045D
07/23/2020	Solid	L3439-14	Corrosivity	Cool 4 deg C	AQUA04	I13	2020444-3002D	07/23/2020	9045D

Date/Time 07/24/2020 16:00

Received by: JH CWT-Intern

Relinquished by: KS (wetchem)

Date/Time 07/24/2020 18:120

Received by: KS (wetchem)

Relinquished by: JH CWT-Intern