



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110202

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/24/2020	17:10
2	CAL2	1	Water	NA	NA	20.3	7.00	07/24/2020	17:11
3	CAL3	1	Water	NA	NA	20.3	10.03	07/24/2020	17:12
4	ICV	1	Water	NA	NA	20.3	7.00	07/24/2020	17:15
5	CCV1	1	Water	NA	NA	20.4	2.02	07/24/2020	17:18
6	L3432-08	1	Water	NA	NA	24.9	6.75	07/24/2020	17:40
7	L3432-10	1	Water	NA	NA	23.2	5.97	07/24/2020	17:41
8	L3432-11	1	Water	NA	NA	23.2	5.17	07/24/2020	17:42
9	L3432-12	1	Water	NA	NA	24.5	7.96	07/24/2020	17:54
10	L3432-13	1	Water	NA	NA	24.7	7.37	07/24/2020	18:00
11	L3432-14	1	Water	NA	NA	24.9	8.53	07/24/2020	18:01
12	L3432-17	1	Water	NA	NA	23.2	8.23	07/24/2020	18:02
13	L3450-02	1	Water	NA	NA	20.8	6.91	07/24/2020	18:05
14	L3450-03	1	Water	NA	NA	21.2	7.19	07/24/2020	18:06
15	L3450-04	1	Water	NA	NA	22.1	7.00	07/24/2020	18:10
16	CCV2	1	Water	NA	NA	20.3	12.02	07/24/2020	18:11
17	L3452-01	1	Water	NA	NA	20.2	7.23	07/24/2020	18:12
18	L3452-02	1	Water	NA	NA	19.9	7.24	07/24/2020	18:15
19	L3452-03	1	Water	NA	NA	19.8	7.35	07/24/2020	18:16
20	L3452-04	1	Water	NA	NA	20.2	7.28	07/24/2020	18:19

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	L3452-05	1	Water	NA	NA	20.3	7.34	07/24/2020	18:20
22	L3452-06	1	Water	NA	NA	20.1	7.27	07/24/2020	18:22
23	L3452-07	1	Water	NA	NA	20.4	7.11	07/24/2020	18:25
24	L3452-08	1	Water	NA	NA	19.3	7.17	07/24/2020	18:26
25	L3452-08DUP	1	Water	NA	NA	19.4	7.18	07/24/2020	18:27
26	CCV3	1	Water	NA	NA	20.4	2.02	07/24/2020	18:30

10202

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13452 WorkList ID : 140810 Department : Wet-Chemistry Date : 07-24-2020 15:52:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/24/2020	Water	L3432-08	pH	Cool 4 deg C	PSEG01	I22	31165-L	07/24/2020	9040C
07/24/2020	Water	L3432-10	pH	Cool 4 deg C	PSEG01	I22	31138	07/24/2020	9040C
07/24/2020	Water	L3432-11	pH	Cool 4 deg C	PSEG01	I22	31176	07/24/2020	9040C
07/24/2020	Water	L3432-12	pH	Cool 4 deg C	PSEG01	I22	31169	07/24/2020	9040C
07/24/2020	Water	L3432-13	pH	Cool 4 deg C	PSEG01	I22	31170	07/24/2020	9040C
07/24/2020	Water	L3432-14	pH	Cool 4 deg C	PSEG01	I22	31171	07/24/2020	9040C
07/24/2020	Water	L3432-17	pH	Cool 4 deg C	PSEG01	I22	69-71	07/24/2020	9040C
07/23/2020	Water	L3450-02	pH	Cool 4 deg C	TETR06	I13	TT-016-IDWGW-20200723	07/23/2020	9040C
07/23/2020	Water	L3450-03	pH	Cool 4 deg C	TETR06	I13	TT-017-IDWGW-20200723	07/23/2020	9040C
07/23/2020	Water	L3450-04	pH	Cool 4 deg C	TETR06	I13	TT-018-IDWGW-20200723	07/23/2020	9040C
07/23/2020	Water	L3452-01	pH	Cool 4 deg C	TETR16	I22	C0AA7	07/23/2020	9040C
07/23/2020	Water	L3452-02	pH	Cool 4 deg C	TETR16	I22	C0AA8	07/23/2020	9040C
07/23/2020	Water	L3452-03	pH	Cool 4 deg C	TETR16	I22	C0AA9	07/23/2020	9040C
07/23/2020	Water	L3452-04	pH	Cool 4 deg C	TETR16	I22	C0AB0	07/23/2020	9040C
07/23/2020	Water	L3452-05	pH	Cool 4 deg C	TETR16	I22	C0AB1	07/23/2020	9040C
07/23/2020	Water	L3452-06	pH	Cool 4 deg C	TETR16	I22	C0AB2	07/23/2020	9040C
07/23/2020	Water	L3452-07	pH	Cool 4 deg C	TETR16	I22	C0AB3	07/23/2020	9040C
07/23/2020	Water	L3452-08	pH	Cool 4 deg C	TETR16	I22	C0AB7	07/23/2020	9040C

Date/Time 07/24/2020 16:10
 Received by: Jr (Wet-Chem)
 Relinquished by: KS (Wet-Chem)

Date/Time 07/24/2020 18:30
 Received by: KS (Wet-Chem)
 Relinquished by: JH (Wet-Chem)