


LB110218

ISM02.4

INSTRUMENT ID: CV4



Linear

A= 0.0000e+000

B= 2.5952e-004

C= -2.4879e-002

Rho= 0.9999623

Accept=Accepted

SPPC
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.010	-0.010	56	0.000	56					
0.05	0.050					247					
0.20	0.200	0.195	-0.005	849	0.0 %	849					
2.50	2.500	2.504	0.004	9744	0.0 %	9744					
5.00	5.000	4.994	-0.006	19338	0.0 %	19338					
7.50	7.500	7.563	0.063	29237	0.0 %	29237					
10.0	10.000	9.955	-0.045	38454	0.0 %	38454					

%D
0-0.001

LB110218
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	56	-		0 ISM02.4	PPB	7/27/2020 15:06	Std
	0.2 S0.2	849	-		0.2 ISM02.4	PPB	7/27/2020 15:08	Std
	2.5 S2.5	9744	-		2.5 ISM02.4	PPB	7/27/2020 15:12	Std
	5 S5.0	19338	-		5 ISM02.4	PPB	7/27/2020 15:15	Std
	7.5 S7.5	29237	-		7.5 ISM02.4	PPB	7/27/2020 15:17	Std
	10 S10.0	38454	-		10 ISM02.4	PPB	7/27/2020 15:19	Std
ICV	ICV	14169	3.6522 -		ISM02.4	PPB	7/27/2020 15:22	SMPL
ICB	ICB	-2	-0.0254 -		ISM02.4	PPB	7/27/2020 15:27	SMPL
CCV077	CCV077	19317	4.9882 -		ISM02.4	PPB	7/27/2020 15:29	SMPL
CCB077	CCB077	-32	-0.0332 -		ISM02.4	PPB	7/27/2020 15:34	SMPL
PB130503BL	PBW503	-47	-0.0371 -		ISM02.4	PPB	7/27/2020 15:36	SMPL
L3353-01	METVB9	5	-0.0236 -		ISM02.4	PPB	7/27/2020 15:38	SMPL
L3353-02	METVC0	-47	-0.0371 -		ISM02.4	PPB	7/27/2020 15:41	SMPL
L3353-03	METVC1	-46	-0.0368 -		ISM02.4	PPB	7/27/2020 15:43	SMPL
L3353-04	METVC2	47	-0.0127 -		ISM02.4	PPB	7/27/2020 15:45	SMPL
L3353-05	METVC3	-45	-0.0366 -		ISM02.4	PPB	7/27/2020 15:47	SMPL
L3353-06	METVC4	-52	-0.0384 -		ISM02.4	PPB	7/27/2020 15:50	SMPL
L3353-07	METVD2	-19	-0.0298 -		ISM02.4	PPB	7/27/2020 15:52	SMPL
L3353-08	METVD3	-18	-0.0296 -		ISM02.4	PPB	7/27/2020 15:54	SMPL
L3353-09	METVD3D	-24	-0.0311 -		ISM02.4	PPB	7/27/2020 15:57	SMPL
L3353-10	METVD3S	4155	1.0534 -		ISM02.4	PPB	7/27/2020 15:59	SMPL
L3353-11	METVD4	-81	-0.0459 -		ISM02.4	PPB	7/27/2020 16:01	SMPL
L3353-12	METVD9	-52	-0.0384 -		ISM02.4	PPB	7/27/2020 16:03	SMPL
L3353-13	METVE0	-39	-0.035 -		ISM02.4	PPB	7/27/2020 16:06	SMPL
L3353-14	METVE4	-60	-0.0405 -		ISM02.4	PPB	7/27/2020 16:08	SMPL
PB130504BL	PBW504	-27	-0.0319 -		ISM02.4	PPB	7/27/2020 16:10	SMPL
L3375-01	METVE6	545	0.1166 -		ISM02.4	PPB	7/27/2020 16:12	SMPL
L3375-02	METVE7	499	0.1046 -		ISM02.4	PPB	7/27/2020 16:15	SMPL
L3375-03	METVE8	1533	0.373 -		ISM02.4	PPB	7/27/2020 16:17	SMPL
L3375-04	METVE9	738	0.1666 -		ISM02.4	PPB	7/27/2020 16:19	SMPL
CCV078	CCV078	18107	4.6742 -		ISM02.4	PPB	7/27/2020 16:21	SMPL
CCB078	CCB078	-205	-0.0781 -		ISM02.4	PPB	7/27/2020 16:24	SMPL
L3375-05	METVF0	-80	-0.0456 -		ISM02.4	PPB	7/27/2020 16:26	SMPL
L3375-06	METVF1	-74	-0.0441 -		ISM02.4	PPB	7/27/2020 16:28	SMPL
L3375-07	METVF2	1718	0.421 -		ISM02.4	PPB	7/27/2020 16:30	SMPL
L3375-08	METVF3	-223	-0.0828 -		ISM02.4	PPB	7/27/2020 16:33	SMPL
L3375-09	METVF4	1342	0.3234 -		ISM02.4	PPB	7/27/2020 16:35	SMPL
L3375-10	METVF6	-114	-0.0545 -		ISM02.4	PPB	7/27/2020 16:37	SMPL
L3375-11	METVF7	-84	-0.0467 -		ISM02.4	PPB	7/27/2020 16:40	SMPL
L3375-12	METVF8	-77	-0.0449 -		ISM02.4	PPB	7/27/2020 16:42	SMPL
L3375-13	METVF8D	-55	-0.0392 -		ISM02.4	PPB	7/27/2020 16:44	SMPL
L3375-14	METVF8S	4113	1.0425 -		ISM02.4	PPB	7/27/2020 16:46	SMPL
PB130505BL	PBW505	-130	-0.0586 -		ISM02.4	PPB	7/27/2020 16:49	SMPL
L3376-01	METYG0	305	0.0543 -		ISM02.4	PPB	7/27/2020 16:51	SMPL
L3376-02	METYG1	197	0.0262 -		ISM02.4	PPB	7/27/2020 16:53	SMPL
L3376-03	METYG2	749	0.1695 -		ISM02.4	PPB	7/27/2020 16:55	SMPL

LB110218

INSTRUMENT ID : CV1

L3376-04	METYG3	400	0.0789 -	ISM02.4	PPB	7/27/2020 16:58 SMPL
L3376-05	METYG4	-104	-0.0519 -	ISM02.4	PPB	7/27/2020 17:00 SMPL
L3376-06	METYG5	-39	-0.035 -	ISM02.4	PPB	7/27/2020 17:02 SMPL
L3376-07	METYG6	651	0.1441 -	ISM02.4	PPB	7/27/2020 17:04 SMPL
L3376-08	METYG7	-117	-0.0552 -	ISM02.4	PPB	7/27/2020 17:07 SMPL
L3376-09	METYG8	607	0.1326 -	ISM02.4	PPB	7/27/2020 17:09 SMPL
CCV079	CCV079	17855	4.6088 -	ISM02.4	PPB	7/27/2020 17:11 SMPL
CCB079	CCB079	-167	-0.0682 -	ISM02.4	PPB	7/27/2020 17:14 SMPL
L3376-10	METYG9	193	0.0252 -	ISM02.4	PPB	7/27/2020 17:16 SMPL
L3376-11	METYH0	106	0.0026 -	ISM02.4	PPB	7/27/2020 17:18 SMPL
L3376-12	METYH1	128	0.0083 -	ISM02.4	PPB	7/27/2020 17:20 SMPL
L3376-13	METYH1D	144	0.0125 -	ISM02.4	PPB	7/27/2020 17:23 SMPL
L3376-14	METYH1S	4776	1.2146 -	ISM02.4	PPB	7/27/2020 17:25 SMPL
PB130506BL	PBW506	173	0.02 -	ISM02.4	PPB	7/27/2020 17:31 SMPL
L3425-01	METL78	799	0.1825 -	ISM02.4	PPB	7/27/2020 17:33 SMPL
L3425-02	METL81	1258	0.3016 -	ISM02.4	PPB	7/27/2020 17:35 SMPL
L3425-03	METL82	710	0.1594 -	ISM02.4	PPB	7/27/2020 17:37 SMPL
L3425-04	METL83	733	0.1653 -	ISM02.4	PPB	7/27/2020 17:40 SMPL
L3425-05	METL86	165	0.0179 -	ISM02.4	PPB	7/27/2020 17:42 SMPL
L3425-06	METL88	146	0.013 -	ISM02.4	PPB	7/27/2020 17:44 SMPL
L3425-07	METL89	207	0.0288 -	ISM02.4	PPB	7/27/2020 17:46 SMPL
L3425-08	METL90	190	0.0244 -	ISM02.4	PPB	7/27/2020 17:49 SMPL
L3425-09	METL90D	209	0.0294 -	ISM02.4	PPB	7/27/2020 17:51 SMPL
L3425-10	METL90S	4450	1.13 -	ISM02.4	PPB	7/27/2020 17:53 SMPL
L3425-11	METL97	151	0.0143 -	ISM02.4	PPB	7/27/2020 17:56 SMPL
L3425-12	METL95	710	0.1594 -	ISM02.4	PPB	7/27/2020 17:58 SMPL
L3425-13	METL96	181	0.0221 -	ISM02.4	PPB	7/27/2020 18:00 SMPL
L3425-14	METL99	96	0 -	ISM02.4	PPB	7/27/2020 18:02 SMPL
CCV080	CCV080	17942	4.6314 -	ISM02.4	PPB	7/27/2020 18:05 SMPL
CCB080	CCB080	-184	-0.0726 -	ISM02.4	PPB	7/27/2020 18:07 SMPL
L3425-15	METLA0	1218	0.2912 -	ISM02.4	PPB	7/27/2020 18:09 SMPL
L3425-16	METLA1	2107	0.5219 -	ISM02.4	PPB	7/27/2020 18:11 SMPL
L3425-17	METLA2	652	0.1443 -	ISM02.4	PPB	7/27/2020 18:14 SMPL
L3425-18	METLA4	735	0.1659 -	ISM02.4	PPB	7/27/2020 18:16 SMPL
L3425-19	METLA5	961	0.2245 -	ISM02.4	PPB	7/27/2020 18:18 SMPL
L3425-20	METYB0	167	0.0185 -	ISM02.4	PPB	7/27/2020 18:21 SMPL
L3425-21	METYB1	192	0.0249 -	ISM02.4	PPB	7/27/2020 18:23 SMPL
L3425-22	METYB2	202	0.0275 -	ISM02.4	PPB	7/27/2020 18:25 SMPL
CCV081	CCV081	17889	4.6177 -	ISM02.4	PPB	7/27/2020 18:27 SMPL
CCB081	CCB081	-127	-0.0578 -	ISM02.4	PPB	7/27/2020 18:30 SMPL