



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110229

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	07/27/2020	17:20
2	CAL2	1	Water	NA	NA	21.4	7.00	07/27/2020	17:21
3	CAL3	1	Water	NA	NA	21.4	10.03	07/27/2020	17:22
4	ICV	1	Water	NA	NA	21.3	6.98	07/27/2020	17:25
5	CCV1	1	Water	NA	NA	21.4	2.01	07/27/2020	17:30
6	L3181-13	1	Solid	20.03	20	24.7	8.44	07/27/2020	17:32
7	L3181-13DUP	1	Solid	20.04	20	24.6	8.45	07/27/2020	17:33
8	L3185-06	1	Solid	20.02	20	25.3	7.95	07/27/2020	17:35
9	CCV2	1	Water	NA	NA	20.3	12.02	07/27/2020	17:36

10229

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3185 WorkList ID : 140862 Department : Wet-Chemistry Date : 07-27-2020 14:49:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/27/2020	Solid	L3181-13	Corrosivity	Cool 4 deg C	PSEG05		CL-01-072720	07/27/2020	9045D
07/27/2020	Solid	L3185-06	Corrosivity	Cool 4 deg C	PSEG05	I31	PL-05-072720	07/27/2020	9045D

Date/Time 07/27/2020 16:00
Received by: Jm (Wet Chem)
Relinquished by: KS (Wet Chem)

Date/Time 07/27/2020 14:35
Received by: KS (Wet Chem)
Relinquished by: Jm (Wet Chem)