


LB110236

ISM02.4

INSTRUMENT ID: CV2



Linear

A= 0.0000e+000

B= 2.2832e-004

C= -7.9634e-002

Rho= 0.9994376

Accept=Accepted

slope
x-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.063	0.063	625	0.000	625					-
0.05	0.050					247					-
0.20	0.200	0.174	-0.026	1111	0.0 %	1111					-13
2.50	2.500	2.423	-0.077	10963	0.0 %	10963					-3
5.00	5.000	4.938	-0.062	21975	0.0 %	21975					-1
7.50	7.500	7.741	0.241	34253	0.0 %	34253					-3
10.0	10.000	9.861	-0.139	43537	0.0 %	43537					-1

LB110236
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	625	-		0 ISM02.4	PPB	7/28/2020 11:00	Std
	0.2 S0.2	1111	-		0.2 ISM02.4	PPB	7/28/2020 11:02	Std
	2.5 S2.5	10963	-		2.5 ISM02.4	PPB	7/28/2020 11:04	Std
	5 S5.0	21975	-		5 ISM02.4	PPB	7/28/2020 11:07	Std
	7.5 S7.5	34253	-		7.5 ISM02.4	PPB	7/28/2020 11:09	Std
	10 S10.0	43537	-		10 ISM02.4	PPB	7/28/2020 11:14	Std
ICV	ICV	16836	3.7644 -		ISM02.4	PPB	7/28/2020 11:17	SMPL
ICB	ICB	414	0.0148 -		ISM02.4	PPB	7/28/2020 11:19	SMPL
CCV087	CCV087	21948	4.9315 -		ISM02.4	PPB	7/28/2020 11:21	SMPL
CCB087	CCB087	320	-0.0066 -		ISM02.4	PPB	7/28/2020 11:24	SMPL
PB130534BL	PBS534	103	-0.0562 -		ISM02.4	PPB	7/28/2020 11:26	SMPL
L3349-01	METYC5	81066	18.4296 -		ISM02.4	PPB	7/28/2020 11:28	SMPL
L3349-02	METYC6	30179	6.8109 -		ISM02.4	PPB	7/28/2020 11:31	SMPL
L3349-03	METYC7	103040	23.4467 -		ISM02.4	PPB	7/28/2020 11:33	SMPL
L3349-04	METYC8	491	0.0324 -		ISM02.4	PPB	7/28/2020 11:38	SMPL
L3349-05	METYC9	2600	0.5139 -		ISM02.4	PPB	7/28/2020 11:40	SMPL
L3349-06	METYD0	3945	0.821 -		ISM02.4	PPB	7/28/2020 11:42	SMPL
L3349-07	METYD5	5277	1.1252 -		ISM02.4	PPB	7/28/2020 11:45	SMPL
L3349-08	METYD6	535	0.0425 -		ISM02.4	PPB	7/28/2020 11:47	SMPL
L3349-09	METYD6D	370	0.0048 -		ISM02.4	PPB	7/28/2020 11:52	SMPL
L3349-10	METYD6S	11945	2.6476 -		ISM02.4	PPB	7/28/2020 11:54	SMPL
L3349-11	METYD7	605	0.0584 -		ISM02.4	PPB	7/28/2020 11:56	SMPL
L3349-12	METYD8	485	0.031 -		ISM02.4	PPB	7/28/2020 11:59	SMPL
L3349-13	METYE1	6541	1.4138 -		ISM02.4	PPB	7/28/2020 12:01	SMPL
L3349-14	METYE2	3932	0.8181 -		ISM02.4	PPB	7/28/2020 12:03	SMPL
CCV088	CCV088	20201	4.5327 -		ISM02.4	PPB	7/28/2020 12:05	SMPL
CCB088	CCB088	42	-0.0701 -		ISM02.4	PPB	7/28/2020 12:08	SMPL
L3349-01DLX5	METYC5	14422	3.2132 -		ISM02.4	PPB	7/28/2020 12:10	SMPL
L3349-03DLX5	METYC7	19280	4.3224 -		ISM02.4	PPB	7/28/2020 12:12	SMPL
CCV089	CCV089	22366	5.027 -		ISM02.4	PPB	7/28/2020 12:16	SMPL
CCB089	CCB089	6	-0.0783 -		ISM02.4	PPB	7/28/2020 12:18	SMPL