



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110237

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	07/28/2020	12:05
2	CAL2	1	Water	NA	NA	21.6	7.00	07/28/2020	12:06
3	CAL3	1	Water	NA	NA	21.6	10.04	07/28/2020	12:07
4	ICV	1	Water	NA	NA	21.7	7.00	07/28/2020	12:09
5	CCV1	1	Water	NA	NA	21.7	2.02	07/28/2020	12:10
6	L3468-01	1	Water	NA	NA	27.7	5.40	07/28/2020	12:11
7	L3468-01DUP	1	Water	NA	NA	27.8	5.41	07/28/2020	12:12
8	CCV2	1	Water	NA	NA	21.6	12.02	07/28/2020	12:13

LB 110234

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 13468 WorkList ID : 140909 Department : Wet-Chemistry Date : 07-28-2020 11:50:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/17/2020	Water	L3468-01	pH	Cool 4 deg C	STER11	I31	POOL-WATER	07/17/2020	SM 4500-H B

Date/Time 07/28/2020 12:00
Received by: Jo (wet-chem)
Relinquished by: Jo (wet-chem)

Date/Time 07/28/2020 13:00
Received by: AM (wet-chem)
Relinquished by: Jo (wet-chem)