



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110239

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/28/2020	10:50
2	L3432-08		1	8	100.00	>212.0	>212.0	07/28/2020	11:20
3	L3432-08DUP		1	8	100.00	>212.0	>212.0	07/28/2020	11:40
4	L3432-15		1	9	100.00	>212.0	>212.0	07/28/2020	12:05
5	L3432-17		1	8	100.00	>212.0	>212.0	07/28/2020	12:35
6	L3457-01		1	8	100.00	>212.0	>212.0	07/28/2020	13:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB110239

WorkList Name : fp72820 WorkList ID : 140904 Department : Wet-Chemistry Date : 07-28-2020 10:17:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2020	Water	L3432-08	Flash Point	Cool 4 deg C	PSEG01	I22	31165-L	07/24/2020	1010A
08/07/2020	Water	L3432-15	Flash Point	Cool 4 deg C	PSEG01	I22	38-76	07/24/2020	1010A
08/07/2020	Water	L3432-17	Flash Point	Cool 4 deg C	PSEG01	I22	69-71	07/24/2020	1010A
07/31/2020	Water	L3457-01	Flash Point	Cool 4 deg C	DVIR01	I21	SHB-MOTOR-ROOM	07/23/2020	1010A

Date/Time 07/28/2020 10:25
 Received by: KS (Webcam)
 Relinquished by: JO (Webcam)

Date/Time 07/28/2020 13:10
 Received by: JR (Webcam)
 Relinquished by: KS (Webcam)