



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110250

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	07/28/2020	14:15
2	CAL2	1	Water	NA	NA	21.6	7.00	07/28/2020	14:16
3	CAL3	1	Water	NA	NA	21.6	10.03	07/28/2020	14:17
4	ICV	1	Water	NA	NA	21.6	6.98	07/28/2020	14:19
5	CCV1	1	Water	NA	NA	21.6	2.01	07/28/2020	14:20
6	L3464-01	1	Solid	20.03	20	26.2	7.93	07/28/2020	14:21
7	L3464-01DUP	1	Solid	20.04	20	26.3	7.94	07/28/2020	14:22
8	L3464-02	1	Solid	20.02	20	25.3	6.20	07/28/2020	14:25
9	CCV2	1	Water	NA	NA	21.7	12.02	07/28/2020	14:30

VB 110250

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L3464 WorkList ID : 140922 Department : Wet-Chemistry Date : 07-28-2020 14:03:56

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/28/2020	Solid	L3464-01	pH	Cool 4 deg C	PSEG01	I33	32537	07/28/2020	9045D
07/28/2020	Solid	L3464-02	pH	Cool 4 deg C	PSEG01	I33	32538	07/28/2020	9045D

Date/Time 07/28/2020 16:10
Received by: JO wet
Relinquished by: JO wet

Date/Time 07/28/2020 15:30
Received by: JO wet
Relinquished by: JO wet