



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110251

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.6	4.01	07/28/2020	14:15
2	CAL2	1	Water	NA	NA	21.6	7.00	07/28/2020	14:16
3	CAL3	1	Water	NA	NA	21.6	10.03	07/28/2020	14:17
4	ICV	1	Water	NA	NA	21.6	6.98	07/28/2020	14:19
5	CCV1	1	Water	NA	NA	21.6	2.01	07/28/2020	14:20
6	L3466-01	1	Water	NA	NA	23.8	5.66	07/28/2020	14:35
7	L3466-01DUP	1	Water	NA	NA	23.9	5.66	07/28/2020	14:36
8	CCV2	1	Water	NA	NA	21.7	12.02	07/28/2020	14:39

VB 110251

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3466 W WorkList ID : 140923 Department : Wet-Chemistry Date : 07-28-2020 14:04:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/28/2020	Water	L3466-01	pH	Cool 4 deg C	PSEG01	I32	32530	07/28/2020	9040C

Date/Time 07/28/2020 14:10
Received by: JO [Signature]
Relinquished by: JO [Signature]

Date/Time 07/28/2020 15:30
Received by: JO [Signature]
Relinquished by: JO [Signature]