



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110257

Slope : 99.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	22.3	4.01	07/28/2020	17:20
2	CAL2	1	Water	NA	NA	22.3	6.98	07/28/2020	17:21
3	CAL3	1	Water	NA	NA	22.4	10.03	07/28/2020	17:22
4	ICV	1	Water	NA	NA	22.4	7.00	07/28/2020	17:25
5	CCV1	1	Water	NA	NA	22.4	2.01	07/28/2020	17:26
6	L3183-22	1	Solid	20.03	20	26.4	8.86	07/28/2020	17:27
7	L3183-22DUP	1	Solid	20.04	20	26.5	8.87	07/28/2020	17:28
8	CCV2	1	Water	NA	NA	21.4	12.03	07/28/2020	17:30

LB 110257

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3164 WorkList ID : 140927 Department : Wet-Chemistry Date : 07-28-2020 15:16:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/28/2020	Solid	L3183-22	Corrosivity	Cool 4 deg C	PSEG05		OK-02-072820	07/28/2020	9045D

Date/Time 07/28/2020 16:30
Received by: JQ (wet) (hpm)
Relinquished by: CP (sm)

Date/Time 07/28/2020 18:00
Received by: CP (sm)
Relinquished by: M (wet) (hpm)