


LB 110268

ISM02.4

INSTRUMENT 1D: CV2



Linear

A= 0.0000e+000

B= 2.4040e-004

C= -4.6676e-002

Rho= 0.9997202

Accept = Accepted

slope
y = 1.0102e-004 x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.025	0.025	300	0.000	300					-
0.05	0.050					294					-
0.20	0.200	0.193	-0.007	999	0.0 %	999					-4
2.50	2.500	2.461	-0.039	10430	0.0 %	10430					-2
5.00	5.000	4.942	-0.058	20753	0.0 %	20753					-1
7.50	7.500	7.675	0.175	32121	0.0 %	32121					2
10.0	10.000	9.903	-0.097	41387	0.0 %	41387					-1

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	300	-		0 ISM02.4	PPB	7/29/2020 14:44	Std
	0.2 S0.2	999	-		0.2 ISM02.4	PPB	7/29/2020 14:46	Std
	2.5 S2.5	10430	-		2.5 ISM02.4	PPB	7/29/2020 14:49	Std
	5 S5.0	20753	-		5 ISM02.4	PPB	7/29/2020 14:51	Std
	7.5 S7.5	32121	-		7.5 ISM02.4	PPB	7/29/2020 14:53	Std
	10 S10.0	41387	-		10 ISM02.4	PPB	7/29/2020 14:55	Std
ICV	ICV	16706	3.9695 -		ISM02.4	PPB	7/29/2020 14:58	SMPL
ICB	ICB	204	0.0024 -		ISM02.4	PPB	7/29/2020 15:00	SMPL
CCV003	CCV003	21381	5.0933 -		ISM02.4	PPB	7/29/2020 15:03	SMPL
CCB003	CCB003	169	-0.006 -		ISM02.4	PPB	7/29/2020 15:05	SMPL
PB130575BL	PBS575	155	-0.0094 -		ISM02.4	PPB	7/29/2020 15:07	SMPL
L3392-01	MC0AD0	2040	0.4437 -		ISM02.4	PPB	7/29/2020 15:10	SMPL
L3392-02	MC0AD1	1169	0.2344 -		ISM02.4	PPB	7/29/2020 15:17	SMPL
L3392-03	MC0AD1D	1138	0.2269 -		ISM02.4	PPB	7/29/2020 15:19	SMPL
L3392-04	MC0AD1S	12562	2.9732 -		ISM02.4	PPB	7/29/2020 15:21	SMPL
L3392-05	MC0AD2	2864	0.6418 -		ISM02.4	PPB	7/29/2020 15:24	SMPL
L3392-06	MC0AD3	3550	0.8067 -		ISM02.4	PPB	7/29/2020 15:26	SMPL
L3392-07	MC0AD4	1615	0.3416 -		ISM02.4	PPB	7/29/2020 15:28	SMPL
L3392-08	MC0AD5	1391	0.2877 -		ISM02.4	PPB	7/29/2020 15:30	SMPL
L3392-09	MC0AD6	1841	0.3959 -		ISM02.4	PPB	7/29/2020 15:33	SMPL
L3392-10	MC0AD7	2411	0.5329 -		ISM02.4	PPB	7/29/2020 15:35	SMPL
L3392-11	MC0AD8	1544	0.3245 -		ISM02.4	PPB	7/29/2020 15:37	SMPL
L3392-12	MC0AD9	1616	0.3418 -		ISM02.4	PPB	7/29/2020 15:39	SMPL
L3392-13	MC0AE0	795	0.1444 -		ISM02.4	PPB	7/29/2020 15:42	SMPL
L3392-14	MC0AE2	536	0.0822 -		ISM02.4	PPB	7/29/2020 15:44	SMPL
L3392-15	MC0AE4	435	0.0579 -		ISM02.4	PPB	7/29/2020 15:46	SMPL
L3392-16	MC0AE5	1038	0.2029 -		ISM02.4	PPB	7/29/2020 15:49	SMPL
L3392-17	MC0AE6	835	0.1541 -		ISM02.4	PPB	7/29/2020 15:51	SMPL
L3392-18	MC0AE7	389	0.0468 -		ISM02.4	PPB	7/29/2020 15:53	SMPL
PB130576BL	PBS576	130	-0.0154 -		ISM02.4	PPB	7/29/2020 15:55	SMPL
CCV004	CCV004	22868	5.4508 -		ISM02.4	PPB	7/29/2020 15:58	SMPL
CCB004	CCB004	183	-0.0027 -		ISM02.4	PPB	7/29/2020 16:00	SMPL
L3393-01	MC0AE3	1069	0.2103 -		ISM02.4	PPB	7/29/2020 16:02	SMPL
L3393-02	MC0AE3D	957	0.1834 -		ISM02.4	PPB	7/29/2020 16:07	SMPL
L3393-03	MC0AE3S	13230	3.1338 -		ISM02.4	PPB	7/29/2020 16:09	SMPL
PB130577BL	PBS577	80	-0.0274 -		ISM02.4	PPB	7/29/2020 16:12	SMPL
L3420-01	MC0AE8	772	0.1389 -		ISM02.4	PPB	7/29/2020 16:14	SMPL
L3420-02	MC0AE9	325	0.0315 -		ISM02.4	PPB	7/29/2020 16:16	SMPL
L3420-03	MC0AF0	580	0.0928 -		ISM02.4	PPB	7/29/2020 16:18	SMPL
L3420-04	MC0AF0D	516	0.0774 -		ISM02.4	PPB	7/29/2020 16:21	SMPL
L3420-05	MC0AF0S	12760	3.0208 -		ISM02.4	PPB	7/29/2020 16:23	SMPL
L3420-06	MC0AF1	2150	0.4702 -		ISM02.4	PPB	7/29/2020 16:25	SMPL
L3420-07	MC0AF2	1489	0.3113 -		ISM02.4	PPB	7/29/2020 16:28	SMPL
L3420-08	MC0AF3	1419	0.2945 -		ISM02.4	PPB	7/29/2020 16:30	SMPL
L3420-09	MC0AF8	505	0.0747 -		ISM02.4	PPB	7/29/2020 16:32	SMPL
L3420-10	MC0AG7	524	0.0793 -		ISM02.4	PPB	7/29/2020 16:34	SMPL

LB110268

INSTRUMENT ID : CV1

PB130578BL	PBS578	126	-0.0164 -	ISM02.4	PPB	7/29/2020 16:37	SMPL
L3426-01	METL79	3573	0.8123 -	ISM02.4	PPB	7/29/2020 16:39	SMPL
L3426-02	METL80	2797	0.6257 -	ISM02.4	PPB	7/29/2020 16:41	SMPL
L3426-03	METL84	14182	3.3627 -	ISM02.4	PPB	7/29/2020 16:43	SMPL
L3426-04	METL85	15770	3.7444 -	ISM02.4	PPB	7/29/2020 16:46	SMPL
L3426-05	METL91	1503	0.3146 -	ISM02.4	PPB	7/29/2020 16:48	SMPL
CCV005	CCV005	22855	5.4477 -	ISM02.4	PPB	7/29/2020 16:50	SMPL
CCB005	CCB005	165	-0.007 -	ISM02.4	PPB	7/29/2020 16:52	SMPL
L3426-06	METL92	9146	2.152 -	ISM02.4	PPB	7/29/2020 16:55	SMPL
L3426-07	METL93	742	0.1317 -	ISM02.4	PPB	7/29/2020 16:57	SMPL
L3426-08	METL93D	813	0.1488 -	ISM02.4	PPB	7/29/2020 16:59	SMPL
L3426-09	METL93S	11677	2.7605 -	ISM02.4	PPB	7/29/2020 17:04	SMPL
L3426-10	METYA7	160749	38.5975 -	ISM02.4	PPB	7/29/2020 17:06	SMPL
L3426-11	METYA8	61638	14.7711 -	ISM02.4	PPB	7/29/2020 17:09	SMPL
L3426-12	METYB3	22862	5.4494 -	ISM02.4	PPB	7/29/2020 17:11	SMPL
L3426-13	METYB4	120917	29.0218 -	ISM02.4	PPB	7/29/2020 17:13	SMPL
L3426-14	METYB5	47836	11.4531 -	ISM02.4	PPB	7/29/2020 17:16	SMPL
CCV006	CCV006	22838	5.4436 -	ISM02.4	PPB	7/29/2020 17:18	SMPL
CCB006	CCB006	75	-0.0286 -	ISM02.4	PPB	7/29/2020 17:20	SMPL
L3426-10DLX10		16840	4.0017 -	ISM02.4	PPB	7/29/2020 17:25	SMPL
L3426-11DLX5		12261	2.9009 -	ISM02.4	PPB	7/29/2020 17:27	SMPL
L3426-13DLX5		24585	5.8636 -	ISM02.4	PPB	7/29/2020 17:29	SMPL
L3426-14DLX2		22876	5.4527 -	ISM02.4	PPB	7/29/2020 17:32	SMPL
CCV007	CCV007	21433	5.1058 -	ISM02.4	PPB	7/29/2020 17:44	SMPL
CCB007	CCB007	162	-0.0077 -	ISM02.4	PPB	7/29/2020 17:46	SMPL