



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110272

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/29/2020	17:00
2	CAL2	1	Water	NA	NA	21.2	7.00	07/29/2020	17:01
3	CAL3	1	Water	NA	NA	21.2	10.03	07/29/2020	17:02
4	ICV	1	Water	NA	NA	21.2	7.02	07/29/2020	17:05
5	CCV1	1	Water	NA	NA	21.4	2.02	07/29/2020	17:06
6	L3484-01	1	Solid	20.02	20	25.2	6.17	07/29/2020	17:10
7	L3484-01DUP	1	Solid	20.03	20	25.3	6.18	07/29/2020	17:11
8	L3484-02	1	Solid	20.01	20	25.0	6.20	07/29/2020	17:12
9	CCV2	1	Water	NA	NA	21.3	12.02	07/29/2020	17:15

183 110272

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3484 S WorkList ID : 140970 Department : Wet-Chemistry Date : 07-29-2020 16:10:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/29/2020	Solid	L3484-01	pH	Cool 4 deg C	PSEG01	I42	30982	07/29/2020	9045D
07/29/2020	Solid	L3484-02	pH	Cool 4 deg C	PSEG01	I42	31156	07/29/2020	9045D

Date/Time 07/29/2020 16:10
Received by: Jo (WLF) (mm)
Relinquished by: KS (webchem)

Date/Time 07/29/2020 18:00
Received by: KS (webchem)
Relinquished by: Jo (WLF) (mm)