



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110273

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	07/29/2020	17:00
2	CAL2	1	Water	NA	NA	21.2	7.00	07/29/2020	17:01
3	CAL3	1	Water	NA	NA	21.2	10.03	07/29/2020	17:02
4	ICV	1	Water	NA	NA	21.2	7.02	07/29/2020	17:05
5	CCV1	1	Water	NA	NA	21.4	2.02	07/29/2020	17:06
6	L3180-04	1	Solid	20.03	20	24.9	7.56	07/29/2020	17:20
7	L3180-04DUP	1	Solid	20.04	20	24.8	7.57	07/29/2020	17:21
8	L3483-02	1	Solid	20.03	20	25.9	8.53	07/29/2020	17:22
9	CCV2	1	Water	NA	NA	21.4	12.02	07/29/2020	17:24

285110273

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L3180 WorkList ID : 140962 Department : Wet-Chemistry Date : 07-29-2020 12:56:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/29/2020	Solid	L3180-04	Corrosivity	Cool 4 deg C	PSEG05	I42	HD-02-072920	07/29/2020	9045D
07/29/2020	Solid	L3483-02	Corrosivity	Cool 4 deg C	PSEG01	I41	S-1	07/29/2020	9045D

Date/Time 07-29-2020 16:10
Received by: JB WLT (Inpm)
Relinquished by: KS (Wetchem)

Date/Time 07/29/2020 18:00
Received by: KS (Wetchem)
Relinquished by: JB WLT (Inpm)