



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110281

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	07/27/2020	09:35
2	CAL2	1	Water	NA	NA	21.1	7.00	07/27/2020	09:36
3	CAL3	1	Water	NA	NA	21.1	10.03	07/27/2020	09:37
4	ICV	1	Water	NA	NA	21.1	7.00	07/27/2020	09:38
5	CCV1	1	Water	NA	NA	21.1	2.02	07/27/2020	09:40
6	L3457-01	1	Water	NA	NA	20.6	7.69	07/27/2020	09:42
7	L3457-01DUP	1	Water	NA	NA	20.7	7.70	07/27/2020	09:43
8	CCV2	1	Water	NA	NA	21.3	12.02	07/27/2020	09:45

LB 110281

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L3457 WorkList ID : 140983 Department : Wet-Chemistry Date : 07-27-2020 9:25:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/23/2020	Water	L3457-01	Corrosivity	Cool 4 deg C	DVIR01	I21	SHB-MOTOR-ROOM	07/23/2020	9040C

Date/Time 07/27/2020 9:30
Received by: Jm (Wet) mm
Relinquished by: Jm (Wet) mm

Date/Time 07/27/2020 10:40
Received by: Jm (Wet)
Relinquished by: Jm (Wet) mm