


L3110282

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.057	0.057	21	0.000	21				
0.2	0.200	0.239	0.039	635	0.0 %	635				
2.5	2.500	2.471	-0.029	8187	0.0 %	8187				
5.0	5.000	4.841	-0.159	16206	0.0 %	16206				
7.5	7.500	7.541	0.041	25343	0.0 %	25343				
10.0	10.000	10.052	0.052	33836	0.0 %	33836				

LB110282 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jul 2020 09:42:37

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	21	-	0.00007470A	PPB	30 Jul 2020 10:16:40	S	Std
0.2 - 1	50.2	635	-	0.20007470A	PPB	30 Jul 2020 10:18:47	S	Std
2.5 - 1	52.5	8187	-	2.50007470A	PPB	30 Jul 2020 10:20:56	S	Std
5.0 - 1	55	16206	-	5.00007470A	PPB	30 Jul 2020 10:23:03	S	Std
7.5 - 1	57.5	25343	-	7.50007470A	PPB	30 Jul 2020 10:25:11	S	Std
10.0 - 1	510	33836	-	10.00007470A	PPB	30 Jul 2020 10:27:17	S	Std
ICV46 - 1	ICV46	12215	3.6612	-7470A	PPB	30 Jul 2020 10:29:55	U	SMPL
ICB46 - 1	ICB46	-268	-0.0284	-7470A	PPB	30 Jul 2020 10:32:01	U	SMPL
CCV08 - 1	CCV08	16258	4.8562	-7470A	PPB	30 Jul 2020 10:34:08	U	SMPL
CCB08 - 1	CCB08	-297	-0.0369	-7470A	PPB	30 Jul 2020 10:36:13	U	SMPL
CRA - 1	CRA	477	0.1918	-7470A	PPB	30 Jul 2020 10:38:21	U	SMPL
HighStd - 1	HighStd	33226	9.8713	-7470A	PPB	30 Jul 2020 10:40:27	U	SMPL
ChkStd - 1	ChkStd	23301	6.9378	-7470A	PPB	30 Jul 2020 10:42:33	U	SMPL
PB130595BL - 1	PBW	-486	-0.0928	-7470A	PPB	30 Jul 2020 10:44:42	U	SMPL
PB130595BS - 1	LCSW	12501	3.7457	-7470A	PPB	30 Jul 2020 10:49:15	U	SMPL
L3183-22 - 1	OK-02-072820	-322	-0.0443	-7470A	PPB	30 Jul 2020 10:51:21	U	SMPL
L3183-22DUP - 1	OK-02-072820DUP	-118	0.0160	-7470A	PPB	30 Jul 2020 10:53:30	U	SMPL
L3183-22MS - 1	OK-02-072820MS	11563	3.4685	-7470A	PPB	30 Jul 2020 10:55:37	U	SMPL
L3183-22MSD - 1	OK-02-072820MSD	13652	4.0859	-7470A	PPB	30 Jul 2020 10:57:45	U	SMPL
L3464-01 - 1	32537	-298	-0.0372	-7470A	PPB	30 Jul 2020 10:59:51	U	SMPL
CCV09 - 1	CCV09	16676	4.9797	-7470A	PPB	30 Jul 2020 11:01:59	U	SMPL
CCB09 - 1	CCB09	-380	-0.0615	-7470A	PPB	30 Jul 2020 11:04:05	U	SMPL
L3464-02 - 1	32538	-122	0.0148	-7470A	PPB	30 Jul 2020 11:06:13	U	SMPL
L3473-05 - 1	TP-2-S	-94	0.0231	-7470A	PPB	30 Jul 2020 11:08:19	U	SMPL
L3473-11 - 1	TP-2-D	-85	0.0257	-7470A	PPB	30 Jul 2020 11:10:26	U	SMPL
PB130595TB - 1	PB130571TB	-178	-0.0018	-7470A	PPB	30 Jul 2020 11:14:39	U	SMPL
L3183-22LX5 - 1		-36	0.0402	-7470A	PPB	30 Jul 2020 11:16:48	U	SMPL
L3183-22A - 1		12292	3.6839	-7470A	PPB	30 Jul 2020 11:18:55	U	SMPL
CCV10 - 1	CCV10	16479	4.9215	-7470A	PPB	30 Jul 2020 11:21:03	U	SMPL
CCB10 - 1	CCB10	-297	-0.0369	-7470A	PPB	30 Jul 2020 11:23:12	U	SMPL