


LB 110286

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5296e-004

C= 1.8279e-002

Rho= 0.9999205

Accept=Accepted

slope
 $y = 1.2512e-002$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.055	0.055	145	0.000	145					
0.05	0.050					294					
0.20	0.200	0.224	0.024	814	0.0 %	814					
2.50	2.500	2.431	-0.069	9539	0.0 %	9539					
5.00	5.000	4.943	-0.057	19468	0.0 %	19468					
7.50	7.500	7.511	0.011	29619	0.0 %	29619					
10.0	10.000	10.036	0.036	39601	0.0 %	39601					

%D
 =
 12
 -3
 -1
 0
 0

LB110286
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	145	-		0 ISM02.4	PPB	7/30/2020 12:18	Std
	0.2 S0.2	814	-		0.2 ISM02.4	PPB	7/30/2020 12:20	Std
	2.5 S2.5	9539	-		2.5 ISM02.4	PPB	7/30/2020 12:23	Std
	5 S5.0	19468	-		5 ISM02.4	PPB	7/30/2020 12:25	Std
	7.5 S7.5	29619	-		7.5 ISM02.4	PPB	7/30/2020 12:30	Std
	10 S10.0	39601	-		10 ISM02.4	PPB	7/30/2020 12:35	Std
ICV	ICV	15270	3.881 -		ISM02.4	PPB	7/30/2020 12:38	SMPL
ICB	ICB	-234	-0.0409 -		ISM02.4	PPB	7/30/2020 12:40	SMPL
CCV011	CCV011	19603	4.9771 -		ISM02.4	PPB	7/30/2020 12:42	SMPL
CCB011	CCB011	-318	-0.0622 -		ISM02.4	PPB	7/30/2020 12:45	SMPL
PB130609BL	PBW609	-108	-0.009 -		ISM02.4	PPB	7/30/2020 12:47	SMPL
L3420-11	MC0AA0	-57	0.0039 -		ISM02.4	PPB	7/30/2020 12:49	SMPL
L3420-12	MC0AA0D	-31	0.0104 -		ISM02.4	PPB	7/30/2020 12:52	SMPL
L3420-13	MC0AA0S	3545	0.915 -		ISM02.4	PPB	7/30/2020 12:54	SMPL
L3420-14	MC0AA5	-79	-0.0017 -		ISM02.4	PPB	7/30/2020 12:56	SMPL
L3420-15	MC0AG5	-44	0.0071 -		ISM02.4	PPB	7/30/2020 12:58	SMPL
PB130610BL	PBW610	-41	0.0079 -		ISM02.4	PPB	7/30/2020 13:01	SMPL
L3427-01	METYB7	-33	0.0099 -		ISM02.4	PPB	7/30/2020 13:03	SMPL
L3427-02	METYB7D	10	0.0208 -		ISM02.4	PPB	7/30/2020 13:05	SMPL
L3427-03	METYB7S	3859	0.9945 -		ISM02.4	PPB	7/30/2020 13:07	SMPL
PB130611BL	PBW611	-69	0.0008 -		ISM02.4	PPB	7/30/2020 13:10	SMPL
L3438-01	METYH2	1608	0.425 -		ISM02.4	PPB	7/30/2020 13:12	SMPL
L3438-02	METYH3	-54	0.0046 -		ISM02.4	PPB	7/30/2020 13:14	SMPL
L3438-03	METYH4	1458	0.3871 -		ISM02.4	PPB	7/30/2020 13:16	SMPL
L3438-04	METYH5	-13	0.015 -		ISM02.4	PPB	7/30/2020 13:19	SMPL
L3438-05	METYN4	-168	-0.0242 -		ISM02.4	PPB	7/30/2020 13:21	SMPL
L3438-06	METYN5	1418	0.377 -		ISM02.4	PPB	7/30/2020 13:23	SMPL
L3438-07	METYN6	3922	1.0104 -		ISM02.4	PPB	7/30/2020 13:26	SMPL
L3438-08	METYN7	536	0.1539 -		ISM02.4	PPB	7/30/2020 13:28	SMPL
L3438-09	METYN7D	499	0.1445 -		ISM02.4	PPB	7/30/2020 13:30	SMPL
CCV012	CCV012	18714	4.7522 -		ISM02.4	PPB	7/30/2020 13:32	SMPL
CCB012	CCB012	-223	-0.0381 -		ISM02.4	PPB	7/30/2020 13:35	SMPL
L3438-10	METYN7S	4803	1.2333 -		ISM02.4	PPB	7/30/2020 13:40	SMPL
PB130612BL	PBW612	-191	-0.03 -		ISM02.4	PPB	7/30/2020 13:42	SMPL
L3441-01	MC0AA7	60	0.0335 -		ISM02.4	PPB	7/30/2020 13:44	SMPL
L3441-02	MC0AA8	53	0.0317 -		ISM02.4	PPB	7/30/2020 13:46	SMPL
L3441-03	MC0AA9	96	0.0426 -		ISM02.4	PPB	7/30/2020 13:49	SMPL
L3441-04	MC0AA9D	66	0.035 -		ISM02.4	PPB	7/30/2020 13:51	SMPL
L3441-05	MC0AA9S	4031	1.038 -		ISM02.4	PPB	7/30/2020 13:53	SMPL
L3441-06	MC0AB0	46	0.0299 -		ISM02.4	PPB	7/30/2020 13:55	SMPL
L3441-07	MC0AB1	45	0.0297 -		ISM02.4	PPB	7/30/2020 13:58	SMPL
L3441-08	MC0AB2	128	0.0507 -		ISM02.4	PPB	7/30/2020 14:00	SMPL
L3441-09	MC0AB3	66	0.035 -		ISM02.4	PPB	7/30/2020 14:02	SMPL
L3441-10	MC0AB7	-10	0.0157 -		ISM02.4	PPB	7/30/2020 14:08	SMPL
CCV013	CCV013	18612	4.7264 -		ISM02.4	PPB	7/30/2020 14:10	SMPL
CCB013	CCB013	-293	-0.0558 -		ISM02.4	PPB	7/30/2020 14:12	SMPL