



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110299

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/30/2020	14:35
2	L3465-01		1	8	100.00	>212.0	>212.0	07/30/2020	15:10
3	L3465-01DUP		1	8	100.00	>212.0	>212.0	07/30/2020	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp73020-1 WorkList ID : 141006 Department : Wet-Chemistry Date : 07-30-2020 11:56:41

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/04/2020	Solid	L3465-01	Flash Point	Cool 4 deg C	PSEG01	I33	20802	07/28/2020	1010A

Date/Time 07/30/2020 12:00
 Received by: KS (wet-chem)
 Relinquished by: SA (wet-chem)

Date/Time 07/30/2020 15:45
 Received by: SA (wet-chem)
 Relinquished by: KS (wet-chem)