


7470A

LB110300
INSTRUMENT ID: CV1 μ Abs.:

33033



Linear

A= 0.0000e+000

B= 2.9933e-004

C= 3.0493e-003

Rho= 0.9996094

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	23	0.000	23				
0.2	0.200	0.192	-0.008	630	0.0 %	630				
2.5	2.500	2.507	0.007	8364	0.0 %	8364				
5.0	5.000	4.898	-0.102	16352	0.0 %	16352				
7.5	7.500	7.703	0.203	25724	0.0 %	25724				
10.0	10.000	9.891	-0.109	33033	0.0 %	33033				

LB110300 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jul 2020 16:53:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7470A	PPB	30 Jul 2020 16:53:38	S	Std
0.2 - 1	50.2	630	-	0.2000	7470A	PPB	30 Jul 2020 16:55:46	S	Std
2.5 - 1	52.5	8364	-	2.5000	7470A	PPB	30 Jul 2020 16:57:54	S	Std
5.0 - 1	55	16352	-	5.0000	7470A	PPB	30 Jul 2020 17:00:01	S	Std
7.5 - 1	57.5	25724	-	7.5000	7470A	PPB	30 Jul 2020 17:02:08	S	Std
10.0 - 1	61.0	33033	-	10.0000	7470A	PPB	30 Jul 2020 17:04:15	S	Std
ICV48 - 1	ICV48	12992	3.8920	-	7470A	PPB	30 Jul 2020 17:06:49	U	SMPL
ICB48 - 1	ICB48	-288	-0.0832	-	7470A	PPB	30 Jul 2020 17:08:54	U	SMPL
CCV19 - 1	CCV19	16375	4.9046	-	7470A	PPB	30 Jul 2020 17:11:02	U	SMPL
CCB19 - 1	CCB19	-290	-0.0838	-	7470A	PPB	30 Jul 2020 17:13:07	U	SMPL
CRA - 1	CRA	522	0.1593	-	7470A	PPB	30 Jul 2020 17:15:14	U	SMPL
HighStd - 1	HighStd	33204	9.9421	-	7470A	PPB	30 Jul 2020 17:17:20	U	SMPL
ChkStd - 1	ChkStd	21728	6.5070	-	7470A	PPB	30 Jul 2020 17:19:26	U	SMPL
PB130641BL - 1	PBW	-135	-0.0374	-	7470A	PPB	30 Jul 2020 17:23:57	U	SMPL
PB130641BS - 1	LCSW	13222	3.9608	-	7470A	PPB	30 Jul 2020 17:26:04	U	SMPL
L3434-04 - 1	11936	340	0.1048	-	7470A	PPB	30 Jul 2020 17:28:11	U	SMPL
CCV20 - 1	CCV20	16279	4.8759	-	7470A	PPB	30 Jul 2020 17:30:18	U	SMPL
CCB20 - 1	CCB20	-135	-0.0374	-	7470A	PPB	30 Jul 2020 17:34:44	U	SMPL