


LB 110305

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	-7	0.000	-7				
0.2	0.200	0.188	-0.012	746	0.0 %	746				
2.5	2.500	2.413	-0.087	9369	0.0 %	9369				
5.0	5.000	5.150	0.150	19976	0.0 %	19976				
7.5	7.500	7.552	0.052	29281	0.0 %	29281				
10.0	10.000	9.904	-0.096	38394	0.0 %	38394				

LB110305 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 31 Jul 2020 09:44:48

Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-7	-	0.00007470A	PPB	31 Jul 2020 10:11:27	S	Std
0.2 - 1	51.2	746	-	0.20007470A	PPB	31 Jul 2020 10:18:13	S	Std
2.5 - 1	52.5	9369	-	2.50007470A	PPB	31 Jul 2020 10:20:21	S	Std
5.0 - 1	55	19976	-	5.00007470A	PPB	31 Jul 2020 10:22:28	S	Std
7.5 - 1	57.5	29281	-	7.50007470A	PPB	31 Jul 2020 10:24:35	S	Std
10.0 - 1	510	38394	-	10.00007470A	PPB	31 Jul 2020 10:26:43	S	Std
ICV49 - 1	ICV49	15551	4.0085	-7470A	PPB	31 Jul 2020 10:29:32	U	SMPL
ICB49 - 1	ICB49	-81	-0.0258	-7470A	PPB	31 Jul 2020 10:34:21	U	SMPL
CCV21 - 1	CCV21	20065	5.1734	-7470A	PPB	31 Jul 2020 10:36:26	U	SMPL
CCB21 - 1	CCB21	-42	-0.0157	-7470A	PPB	31 Jul 2020 10:40:54	U	SMPL
CRA - 1	CRA	715	0.1797	-7470A	PPB	31 Jul 2020 10:42:59	U	SMPL
HighStd - 1	HighStd	38649	9.9695	-7470A	PPB	31 Jul 2020 10:45:05	U	SMPL
ChkStd - 1	ChkStd	26572	6.8527	-7470A	PPB	31 Jul 2020 10:47:11	U	SMPL
PB130637BL - 1	PBW	28	0.0024	-7470A	PPB	31 Jul 2020 10:51:51	U	SMPL
PB130637BS - 1	LCSW	14467	3.7287	-7470A	PPB	31 Jul 2020 10:53:58	U	SMPL
L3180-04 - 1	HD-02-072920	-315	-0.0861	-7470A	PPB	31 Jul 2020 10:56:05	U	SMPL
L3180-04DUP - 1	HD-02-072920DUP	-94	-0.0291	-7470A	PPB	31 Jul 2020 10:58:13	U	SMPL
L3180-04MS - 1	HD-02-072920MS	14077	3.6281	-7470A	PPB	31 Jul 2020 11:00:21	U	SMPL
L3180-04MSD - 1	HD-02-072920MSD	13349	3.4402	-7470A	PPB	31 Jul 2020 11:02:29	U	SMPL
L3255-14 - 1	CHRT-57	-284	-0.0781	-7470A	PPB	31 Jul 2020 11:04:34	U	SMPL
CCV22 - 1	CCV22	20502	5.2862	-7470A	PPB	31 Jul 2020 11:06:43	U	SMPL
CCB22 - 1	CCB22	22	0.0008	-7470A	PPB	31 Jul 2020 11:11:17	U	SMPL
L3483-02 - 1	S-1	-90	-0.0281	-7470A	PPB	31 Jul 2020 11:13:23	U	SMPL
L3484-03 - 1	82-56	-60	-0.0203	-7470A	PPB	31 Jul 2020 11:15:29	U	SMPL
L3491-04 - 1	TP-1-S	-36	-0.0141	-7470A	PPB	31 Jul 2020 11:17:35	U	SMPL
L3491-09 - 1	TP-1-D	-40	-0.0152	-7470A	PPB	31 Jul 2020 11:19:42	U	SMPL
L3491-13 - 1	TP-3-S	-46	-0.0167	-7470A	PPB	31 Jul 2020 11:21:48	U	SMPL
PB130638BL - 1	PBW	-39	-0.0149	-7470A	PPB	31 Jul 2020 11:23:55	U	SMPL
PB130638BS - 1	LCSW	14394	3.7099	-7470A	PPB	31 Jul 2020 11:26:02	U	SMPL
L3478-03 - 1	OILY-WATER	-199	-0.0562	-7470A	PPB	31 Jul 2020 11:28:10	U	SMPL
L3478-03DUP - 1	OILY-WATERDUP	38	0.0050	-7470A	PPB	31 Jul 2020 11:30:18	U	SMPL
L3478-03MS - 1	OILY-WATERMS	15282	3.9390	-7470A	PPB	31 Jul 2020 11:32:25	U	SMPL
CCV23 - 1	CCV23	20618	5.3161	-7470A	PPB	31 Jul 2020 11:34:31	U	SMPL
CCB23 - 1	CCB23	11	-0.0020	-7470A	PPB	31 Jul 2020 11:39:19	U	SMPL
L3478-03MSD - 1	OILY-WATERMSD	14923	3.8464	-7470A	PPB	31 Jul 2020 11:41:25	U	SMPL
PB130651BL - 1	PBW	46	0.0070	-7470A	PPB	31 Jul 2020 11:48:23	U	SMPL
PB130651BS - 1	LCSW	13634	3.5137	-7470A	PPB	31 Jul 2020 11:50:29	U	SMPL
L3216-07 - 1	LOD-MDL-WATER-01-QT3-2020	139	0.0310	-7470A	PPB	31 Jul 2020 12:00:09	U	SMPL
L3216-08 - 1	LOQ-WATER-02-QT3-2020	734	0.1846	-7470A	PPB	31 Jul 2020 12:04:42	U	SMPL
L3489-15 - 1	FB	-45	-0.0165	-7470A	PPB	31 Jul 2020 12:06:49	U	SMPL
CCV24 - 1	CCV24	20315	5.2379	-7470A	PPB	31 Jul 2020 12:08:57	U	SMPL
CCB24 - 1	CCB24	-38	-0.0147	-7470A	PPB	31 Jul 2020 12:13:31	U	SMPL
PB130637TB - 1	PB130597TB	-44	-0.0162	-7470A	PPB	31 Jul 2020 12:15:39	U	SMPL
PB130638TB - 1	PB130623TB	-50	-0.0178	-7470A	PPB	31 Jul 2020 12:17:45	U	SMPL
L3180-04LX5 - 1		-30	-0.0126	-7470A	PPB	31 Jul 2020 12:19:51	U	SMPL
L3180-04A - 1		15618	4.0258	-7470A	PPB	31 Jul 2020 12:21:58	U	SMPL
L3478-03LX5 - 1		-260	-0.0719	-7470A	PPB	31 Jul 2020 12:24:04	U	SMPL
L3478-03A - 1		15990	4.1218	-7470A	PPB	31 Jul 2020 12:26:11	U	SMPL
CCV25 - 1	CCV25	20308	5.2361	-7470A	PPB	31 Jul 2020 12:28:18	U	SMPL
CCB25 - 1	CCB25	74	0.0143	-7470A	PPB	31 Jul 2020 13:50:51	U	SMPL