



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110310

Slope : 97.6

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	07/31/2020	08:00
2	CAL2	1	Water	NA	NA	20.3	7.00	07/31/2020	08:01
3	CAL3	1	Water	NA	NA	20.3	10.03	07/31/2020	08:02
4	ICV	1	Water	NA	NA	20.3	7.01	07/31/2020	08:05
5	CCV1	1	Water	NA	NA	20.3	2.02	07/31/2020	08:06
6	L3512-05	1	Water	NA	NA	24.9	7.84	07/31/2020	08:10
7	L3512-05DUP	1	Water	NA	NA	24.8	7.85	07/31/2020	08:11
8	CCV2	1	Water	NA	NA	20.3	12.02	07/31/2020	08:12

WORKLIST(Hardcopy Internal Chain)

18110310

WorkList Name : PH L3512 W

WorkList ID : 141049

Department : Wet-Chemistry

Date : 07-31-2020 7:40:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2020	Water	L3512-05	pH	Cool 4 deg C	PSEG01	I51	USED-ANTIFREEZE-1456	07/30/2020	9040C

Date/Time 07/31/2020 7:50
 Received by: Ja Ja / wet
 Relinquished by: Ja Ja / wet

Date/Time 07/31/2020 9:00
 Received by: Ja Ja / wet
 Relinquished by: Ja Ja / wet