


LB110312

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.6049e-004

C= 6.1569e-003

Rho= 0.9999891

Accept=Accepted

Slope
Y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.005	0.005	-3	0.000	-3					
0.05	0.050					294					
0.20	0.200	0.206	0.006	769	0.0 %	769					
2.50	2.500	2.482	-0.018	9503	0.0 %	9503					
5.00	5.000	4.988	-0.012	19127	0.0 %	19127					
7.50	7.500	7.532	0.032	28892	0.0 %	28892					
10.0	10.000	9.986	-0.014	38312	0.0 %	38312					

%D
11
3
1
00
00

LB110312
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-3	-		0 ISM02.4	PPB	7/31/2020 14:07	Std
	0.2 S0.2	769	-		0.2 ISM02.4	PPB	7/31/2020 14:09	Std
	2.5 S2.5	9503	-		2.5 ISM02.4	PPB	7/31/2020 14:11	Std
	5 S5	19127	-		5 ISM02.4	PPB	7/31/2020 14:15	Std
	7.5 S7.5	28892	-		7.5 ISM02.4	PPB	7/31/2020 14:17	Std
	10 S10	38312	-		10 ISM02.4	PPB	7/31/2020 14:20	Std
ICV	ICV	14843	3.8726	-	ISM02.4	PPB	7/31/2020 14:23	SMPL
ICB	ICB	-6	0.0046	-	ISM02.4	PPB	7/31/2020 14:27	SMPL
CCV026	CCV026	19166	4.9987	-	ISM02.4	PPB	7/31/2020 14:30	SMPL
CCB026	CCB026	35	0.0153	-	ISM02.4	PPB	7/31/2020 14:34	SMPL
PB130642BL	PBW642	-70	-0.0121	-	ISM02.4	PPB	7/31/2020 14:37	SMPL
L3421-02	FIELD-BLANK	-52	-0.0074	-	ISM02.4	PPB	7/31/2020 14:39	SMPL
L3421-03	SY-6	1479	0.3914	-	ISM02.4	PPB	7/31/2020 14:41	SMPL
L3421-04	SY-3	1486	0.3932	-	ISM02.4	PPB	7/31/2020 14:48	SMPL
L3421-04D	SY-3D	1490	0.3943	-	ISM02.4	PPB	7/31/2020 14:58	SMPL
L3421-04S	SY-3S	5882	1.5383	-	ISM02.4	PPB	7/31/2020 15:00	SMPL
L3421-05	SY-3D	793	0.2127	-	ISM02.4	PPB	7/31/2020 15:02	SMPL
L3421-06	SY-3DD	-139	-0.0301	-	ISM02.4	PPB	7/31/2020 15:05	SMPL
L3421-07	SY-7	-18	0.0015	-	ISM02.4	PPB	7/31/2020 15:07	SMPL
L3421-08	FIELD-BLANK	-4	0.0051	-	ISM02.4	PPB	7/31/2020 15:09	SMPL
L3421-09	SY-6	8	0.0082	-	ISM02.4	PPB	7/31/2020 15:12	SMPL
L3421-10	SY-3	1029	0.2742	-	ISM02.4	PPB	7/31/2020 15:14	SMPL
L3421-11	SY-3D	304	0.0853	-	ISM02.4	PPB	7/31/2020 15:16	SMPL
L3421-12	SY-3DD	-102	-0.0204	-	ISM02.4	PPB	7/31/2020 15:18	SMPL
L3421-13	SY-7	-18	0.0015	-	ISM02.4	PPB	7/31/2020 15:21	SMPL
CCV027	CCV027	19287	5.0302	-	ISM02.4	PPB	7/31/2020 15:23	SMPL
CCB027	CCB027	49	0.0189	-	ISM02.4	PPB	7/31/2020 15:25	SMPL
L3442-01	SY-2R	-24	-0.0001	-	ISM02.4	PPB	7/31/2020 15:27	SMPL
L3442-02	SY-2D	9	0.0085	-	ISM02.4	PPB	7/31/2020 15:30	SMPL
L3442-03	SY-2R	-11	0.0033	-	ISM02.4	PPB	7/31/2020 15:32	SMPL
L3442-04	SY-2D	8	0.0082	-	ISM02.4	PPB	7/31/2020 15:34	SMPL
CCV028	CCV028	19125	4.988	-	ISM02.4	PPB	7/31/2020 15:36	SMPL
CCB028	CCB028	-109	-0.0222	-	ISM02.4	PPB	7/31/2020 15:39	SMPL