



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110320

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	07/31/2020	16:45
2	L3512-05		1	8	100.00	>212.0	>212.0	07/31/2020	17:10
3	L3537-04		1	7	100.00	>212.0	>212.0	07/31/2020	17:35
4	L3537-04DUP		1	7	100.00	>212.0	>212.0	07/31/2020	17:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP73120 WorkList ID : 141058 Department : Wet-Chemistry Date : 07-31-2020 16:18:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2020	Water	L3512-05	Flash Point	Cool 4 deg C	PSEG01	I51	USED-ANTIFREEZE-1456	07/30/2020	1010A
08/13/2020	Water	L3537-04	Flash Point	Cool 4 deg C	PARS15	I53	JTY1-WC-L-01	07/30/2020	1010A

Date/Time 07/31/2020 16:20
 Received by: KS (Wetchem)
 Relinquished by: JA (Wetchem)

Date/Time 07/31/2020 18:00
 Received by: JA (Wetchem)
 Relinquished by: KS (Wetchem)