


LB110335

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-63	0.000	-63				
0.2	0.200	0.228	0.028	726	0.0 %	726				
2.5	2.500	2.620	0.120	9051	0.0 %	9051				
5.0	5.000	4.946	-0.054	17148	0.0 %	17148				
7.5	7.500	7.195	-0.305	24974	0.0 %	24974				
10.0	10.000	10.209	0.209	35464	0.0 %	35464				

LB110335 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 03 Aug 2020 09:11:01						
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type	
0.0 - 1	50	-63	-	0.00007470A	PPB		03 Aug 2020 09:30:29	S	Std	
0.2 - 1	50.2	726	-	0.20007470A	PPB		03 Aug 2020 09:35:01	S	Std	
2.5 - 1	52.5	9051	-	2.50007470A	PPB		03 Aug 2020 09:37:09	S	Std	
5.0 - 1	55	17148	-	5.00007470A	PPB		03 Aug 2020 09:39:16	S	Std	
7.5 - 1	57.5	24974	-	7.50007470A	PPB		03 Aug 2020 09:41:24	S	Std	
10.0 - 1	510	35464	-	10.00007470A	PPB		03 Aug 2020 09:48:55	S	Std	
ICV51 - 1	ICV51	13378	3.8632	-7470A	PPB		03 Aug 2020 09:51:30	U	SMPL	
ICB51 - 1	ICB51	-66	0.0003	-7470A	PPB		03 Aug 2020 09:55:57	U	SMPL	
CCV33 - 1	CCV33	17673	5.0973	-7470A	PPB		03 Aug 2020 09:58:03	U	SMPL	
CCB33 - 1	CCB33	-18	0.0141	-7470A	PPB		03 Aug 2020 10:02:32	U	SMPL	
CRA - 1	CRA	751	0.2351	-7470A	PPB		03 Aug 2020 10:04:37	U	SMPL	
HighStd - 1	HighStd	35702	10.2777	-7470A	PPB		03 Aug 2020 10:06:43	U	SMPL	
ChkStd - 1	ChkStd	23856	6.8739	-7470A	PPB		03 Aug 2020 10:08:49	U	SMPL	
PB130668BL - 1	PBW	180	0.0710	-7470A	PPB		03 Aug 2020 10:10:58	U	SMPL	
PB130668BS - 1	LCSW	12864	3.7156	-7470A	PPB		03 Aug 2020 10:13:07	U	SMPL	
L3507-07 - 1	SU-01-073020	-26	0.0118	-7470A	PPB		03 Aug 2020 10:15:14	U	SMPL	
L3507-07DUP - 1	SU-01-073020DUP	17	0.0241	-7470A	PPB		03 Aug 2020 10:17:22	U	SMPL	
L3507-07MS - 1	SU-01-073020MS	11988	3.4638	-7470A	PPB		03 Aug 2020 10:19:29	U	SMPL	
L3507-07MSD - 1	SU-01-073020MSD	13685	3.9515	-7470A	PPB		03 Aug 2020 10:21:37	U	SMPL	
L3511-03 - 1	TP-1	564	0.1813	-7470A	PPB		03 Aug 2020 10:23:44	U	SMPL	
CCV34 - 1	CCV34	17851	5.1485	-7470A	PPB		03 Aug 2020 10:28:14	U	SMPL	
CCB34 - 1	CCB34	-215	-0.0425	-7470A	PPB		03 Aug 2020 10:30:19	U	SMPL	
L3511-08 - 1	TP-2	810	0.2520	-7470A	PPB		03 Aug 2020 10:32:27	U	SMPL	
L3511-13 - 1	TP-3	785	0.2448	-7470A	PPB		03 Aug 2020 10:34:33	U	SMPL	
L3511-18 - 1	SP	840	0.2606	-7470A	PPB		03 Aug 2020 10:36:39	U	SMPL	
L3518-03 - 1	TP-4	918	0.2830	-7470A	PPB		03 Aug 2020 10:38:46	U	SMPL	
L3518-08 - 1	TP-5	675	0.2132	-7470A	PPB		03 Aug 2020 10:40:52	U	SMPL	
L3518-13 - 1	TP-6	653	0.2069	-7470A	PPB		03 Aug 2020 10:42:59	U	SMPL	
L3518-18 - 1	TP-7	768	0.2399	-7470A	PPB		03 Aug 2020 10:45:06	U	SMPL	
L3518-23 - 1	TP-8	782	0.2440	-7470A	PPB		03 Aug 2020 10:47:14	U	SMPL	
L3518-28 - 1	TP-9	673	0.2126	-7470A	PPB		03 Aug 2020 10:49:21	U	SMPL	
L3518-33 - 1	TP-10	700	0.2204	-7470A	PPB		03 Aug 2020 10:51:28	U	SMPL	
CCV35 - 1	CCV35	17986	5.1873	-7470A	PPB		03 Aug 2020 10:53:34	U	SMPL	
CCB35 - 1	CCB35	-72	-0.0014	-7470A	PPB		03 Aug 2020 10:55:40	U	SMPL	
L3518-38 - 1	TP-11	36	0.0296	-7470A	PPB		03 Aug 2020 10:57:48	U	SMPL	
L3518-43 - 1	TP-12	73	0.0402	-7470A	PPB		03 Aug 2020 10:59:54	U	SMPL	
L3518-48 - 1	TP-13	79	0.0420	-7470A	PPB		03 Aug 2020 11:02:01	U	SMPL	
PB130669BL - 1	PBW	72	0.0400	-7470A	PPB		03 Aug 2020 11:04:07	U	SMPL	
PB130669BS - 1	LCSW	13751	3.9704	-7470A	PPB		03 Aug 2020 11:06:13	U	SMPL	
L3216-09 - 1	MDL-WATER-03-QT3-2020	149	0.0621	-7470A	PPB		03 Aug 2020 11:14:01	U	SMPL	
L3523-01 - 1	20200447-FIELD-BLANK	81	0.0425	-7470A	PPB		03 Aug 2020 11:16:08	U	SMPL	
L3523-02 - 1	20200448-SITE-WATER	210	0.0796	-7470A	PPB		03 Aug 2020 11:18:16	U	SMPL	
L3523-02DUP - 1	20200448-SITE-WATERDUP	200	0.0767	-7470A	PPB		03 Aug 2020 11:20:23	U	SMPL	
L3523-02MS - 1	20200448-SITE-WATERMS	14067	4.0612	-7470A	PPB		03 Aug 2020 11:22:30	U	SMPL	
CCV36 - 1	CCV36	17630	5.0850	-7470A	PPB		03 Aug 2020 11:24:36	U	SMPL	
CCB36 - 1	CCB36	2	0.0198	-7470A	PPB		03 Aug 2020 11:29:15	U	SMPL	
L3523-02MSD - 1	20200448-SITE-WATERMSD	14125	4.0779	-7470A	PPB		03 Aug 2020 11:31:21	U	SMPL	
PB130671BL - 1	PBW	-114	-0.0135	-7470A	PPB		03 Aug 2020 11:33:27	U	SMPL	
PB130671BS - 1	LCSW	12911	3.7291	-7470A	PPB		03 Aug 2020 11:35:34	U	SMPL	
L3512-04 - 1	OILY-WATER-1493	-97	-0.0086	-7470A	PPB		03 Aug 2020 11:37:41	U	SMPL	
L3512-04DUP - 1	OILY-WATER-1493DUP	59	0.0362	-7470A	PPB		03 Aug 2020 11:42:30	U	SMPL	
L3512-04MS - 1	OILY-WATER-1493MS	12334	3.5633	-7470A	PPB		03 Aug 2020 11:44:37	U	SMPL	
L3512-04MSD - 1	OILY-WATER-1493MSD	13063	3.7727	-7470A	PPB		03 Aug 2020 11:46:44	U	SMPL	
L3512-05 - 1	USED-ANTIFREEZE-1456	-88	-0.0060	-7470A	PPB		03 Aug 2020 11:48:52	U	SMPL	
PB130668TB - 1	PB130639TB	258	0.0934	-7470A	PPB		03 Aug 2020 11:57:28	U	SMPL	
PB130671TB - 1	PB130670TB	242	0.0888	-7470A	PPB		03 Aug 2020 12:01:56	U	SMPL	
CCV37 - 1	CCV37	18026	5.1988	-7470A	PPB		03 Aug 2020 12:04:02	U	SMPL	
CCB37 - 1	CCB37	-146	-0.0227	-7470A	PPB		03 Aug 2020 12:06:08	U	SMPL	
L3507-07LX5 - 1		41	0.0310	-7470A	PPB		03 Aug 2020 12:13:44	U	SMPL	
L3507-07A - 1		13408	3.8719	-7470A	PPB		03 Aug 2020 12:15:50	U	SMPL	
L3523-02LX5 - 1		-126	-0.0169	-7470A	PPB		03 Aug 2020 12:17:57	U	SMPL	
L3523-02A - 1		13556	3.9144	-7470A	PPB		03 Aug 2020 12:20:04	U	SMPL	
L3512-04LX5 - 1		-176	-0.0313	-7470A	PPB		03 Aug 2020 12:22:11	U	SMPL	
L3512-04A - 1		14177	4.0928	-7470A	PPB		03 Aug 2020 12:24:18	U	SMPL	
CCV38 - 1	CCV38	17699	5.1048	-7470A	PPB		03 Aug 2020 12:28:33	U	SMPL	
CCB38 - 1	CCB38	-24	0.0124	-7470A	PPB		03 Aug 2020 12:30:40	U	SMPL	