


7471B

LB110339
INSTRUMENT ID: CY1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	131	0.000	131				
0.2	0.200	0.162	-0.038	823	0.0 %	823				
2.5	2.500	2.462	-0.038	12719	0.0 %	12719				
5.0	5.000	5.053	0.053	26118	0.0 %	26118				
7.5	7.500	7.546	0.046	39014	0.0 %	39014				
10.0	10.000	9.949	-0.051	51443	0.0 %	51443				

LB110339 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Aug 2020 13:59:35

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	131	-	0.0000	7471B	PPB	03 Aug 2020 14:01:29	S	Std
0.2 - 1	50.2	823	-	0.2000	7471B	PPB	03 Aug 2020 14:03:46	S	Std
2.5 - 1	52.5	12719	-	2.5000	7471B	PPB	03 Aug 2020 14:06:03	S	Std
5.0 - 1	55	26118	-	5.0000	7471B	PPB	03 Aug 2020 14:08:20	S	Std
7.5 - 1	57.5	39014	-	7.5000	7471B	PPB	03 Aug 2020 14:10:37	S	Std
10.0 - 1	510	51443	-	10.0000	7471B	PPB	03 Aug 2020 14:15:31	S	Std
ICV52 - 1	ICV52	20194	3.9073	-	7471B	PPB	03 Aug 2020 14:18:13	U	SMPL
ICB52 - 1	ICB52	-178	-0.0311	-	7471B	PPB	03 Aug 2020 14:20:29	U	SMPL
CCV39 - 1	CCV39	26122	5.0533	-	7471B	PPB	03 Aug 2020 14:22:47	U	SMPL
CCB39 - 1	CCB39	-197	-0.0348	-	7471B	PPB	03 Aug 2020 14:25:01	U	SMPL
CRA - 1	CRA	889	0.1752	-	7471B	PPB	03 Aug 2020 14:30:26	U	SMPL
HighStd - 1	HighStd	51853	10.0278	-	7471B	PPB	03 Aug 2020 14:32:41	U	SMPL
ChkStd - 1	ChkStd	37364	7.2267	-	7471B	PPB	03 Aug 2020 14:34:56	U	SMPL
PB130688BL - 1	PBS	18	0.0068	-	7471B	PPB	03 Aug 2020 14:39:46	U	SMPL
PB130688BS - 1	LCSS	17715	3.4281	-	7471B	PPB	03 Aug 2020 14:42:02	U	SMPL
L3521-03 - 1	COMP-AB	13164	2.5482	-	7471B	PPB	03 Aug 2020 14:44:18	U	SMPL
L3521-07 - 1	COMP-CD	6173	1.1967	-	7471B	PPB	03 Aug 2020 14:46:36	U	SMPL
L3521-07DUP - 1	COMP-CDDUP	6586	1.2765	-	7471B	PPB	03 Aug 2020 14:48:55	U	SMPL
L3521-07MS - 1	COMP-CDMS	25738	4.9791	-	7471B	PPB	03 Aug 2020 14:51:14	U	SMPL
L3521-07MSD - 1	COMP-CDMSD	24367	4.7141	-	7471B	PPB	03 Aug 2020 14:53:30	U	SMPL
CCV40 - 1	CCV40	24916	4.8202	-	7471B	PPB	03 Aug 2020 14:55:47	U	SMPL
CCB40 - 1	CCB40	4	0.0041	-	7471B	PPB	03 Aug 2020 15:00:18	U	SMPL
L3525-01 - 1	TILCON-FL-002	5097	0.9887	-	7471B	PPB	03 Aug 2020 15:02:33	U	SMPL
L3535-01 - 1	JTY1-SB-BL-IB	6640	1.2870	-	7471B	PPB	03 Aug 2020 15:04:48	U	SMPL
L3535-02 - 1	JTY1-SB-0203	5378	1.0430	-	7471B	PPB	03 Aug 2020 15:07:05	U	SMPL
L3535-03 - 1	JTY1-SB-01	2660	0.5175	-	7471B	PPB	03 Aug 2020 15:09:21	U	SMPL
L3535-04 - 1	JTY1-SB-101	2963	0.5761	-	7471B	PPB	03 Aug 2020 15:11:37	U	SMPL
L3538-01 - 1	WASTE	5798	1.1242	-	7471B	PPB	03 Aug 2020 15:13:54	U	SMPL
L3521-07LX5 - 1		789	0.1558	-	7471B	PPB	03 Aug 2020 15:23:06	U	SMPL
L3521-07A - 1		29037	5.6169	-	7471B	PPB	03 Aug 2020 15:25:22	U	SMPL
CCV41 - 1	CCV41	26194	5.0673	-	7471B	PPB	03 Aug 2020 15:27:39	U	SMPL
CCB41 - 1	CCB41	-173	-0.0302	-	7471B	PPB	03 Aug 2020 15:29:57	U	SMPL