


LB110350

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.053	0.053	-24	0.000	-24				
0.2	0.200	0.180	-0.020	426	0.0 %	426				
2.5	2.500	2.511	0.011	8726	0.0 %	8726				
5.0	5.000	4.920	-0.080	17302	0.0 %	17302				
7.5	7.500	7.495	-0.005	26466	0.0 %	26466				
10.0	10.000	10.040	0.040	35529	0.0 %	35529				

LB110350 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 04 Aug 2020 09:57:05				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.00007470A	PPB	04 Aug 2020 10:00:36	S	Std
0.2 - 1	50.2	426	-	0.20007470A	PPB	04 Aug 2020 10:02:44	S	Std
2.5 - 1	52.5	8726	-	2.50007470A	PPB	04 Aug 2020 10:04:52	S	Std
5.0 - 1	55	17302	-	5.00007470A	PPB	04 Aug 2020 10:07:00	S	Std
7.5 - 1	57.5	26466	-	7.50007470A	PPB	04 Aug 2020 10:09:07	S	Std
10.0 - 1	59	35529	-	10.00007470A	PPB	04 Aug 2020 10:11:13	S	Std
ICV54 - 1	ICV54	12853	3.6706	-7470A	PPB	04 Aug 2020 10:13:53	U	SMPL
ICB54 - 1	ICB54	-315	-0.0282	-7470A	PPB	04 Aug 2020 10:15:59	U	SMPL
CCV44 - 1	CCV44	17780	5.0546	-7470A	PPB	04 Aug 2020 10:18:06	U	SMPL
CCB44 - 1	CCB44	-339	-0.0350	-7470A	PPB	04 Aug 2020 10:20:11	U	SMPL
CRA - 1	CRA	486	0.1968	-7470A	PPB	04 Aug 2020 10:36:31	U	SMPL
HighStd - 1	HighStd	36243	10.2409	-7470A	PPB	04 Aug 2020 10:38:37	U	SMPL
ChkStd - 1	ChkStd	25496	7.2221	-7470A	PPB	04 Aug 2020 10:40:43	U	SMPL
PB130707BL - 1	PBW	-129	0.0240	-7470A	PPB	04 Aug 2020 10:45:14	U	SMPL
PB130707BS - 1	LCSW	13081	3.7347	-7470A	PPB	04 Aug 2020 10:47:21	U	SMPL
L3512-01 - 1	OILY-DEBRIS-1499	-148	0.0187	-7470A	PPB	04 Aug 2020 10:49:28	U	SMPL
L3512-02 - 1	ROLLOFF-72-11982	-140	0.0209	-7470A	PPB	04 Aug 2020 10:51:35	U	SMPL
L3521-08 - 1	COMP-CD	-178	0.0102	-7470A	PPB	04 Aug 2020 10:55:50	U	SMPL
L3537-01 - 1	JTY1-WC-S-BL-IB	-31	0.0515	-7470A	PPB	04 Aug 2020 10:57:57	U	SMPL
CCV45 - 1	CCV45	17450	4.9619	-7470A	PPB	04 Aug 2020 11:00:03	U	SMPL
CCB45 - 1	CCB45	-257	-0.0120	-7470A	PPB	04 Aug 2020 11:02:08	U	SMPL
L3537-01DUP - 1	JTY1-WC-S-BL-IBDUP	-68	0.0411	-7470A	PPB	04 Aug 2020 11:04:16	U	SMPL
L3537-01MS - 1	JTY1-WC-S-BL-IBMS	13322	3.8024	-7470A	PPB	04 Aug 2020 11:06:22	U	SMPL
L3537-01MSD - 1	JTY1-WC-S-BL-IBMSD	12765	3.6459	-7470A	PPB	04 Aug 2020 11:08:28	U	SMPL
L3537-02 - 1	JTY1-WC-S-0203	-296	-0.0229	-7470A	PPB	04 Aug 2020 11:10:35	U	SMPL
L3537-03 - 1	JTY1-WC-S-01	-94	0.0338	-7470A	PPB	04 Aug 2020 11:12:43	U	SMPL
PB130708BL - 1	PBW	-56	0.0445	-7470A	PPB	04 Aug 2020 11:14:50	U	SMPL
PB130708BS - 1	LCSW	12434	3.5529	-7470A	PPB	04 Aug 2020 11:16:58	U	SMPL
L3525-02 - 1	TILCON-FL-002	1408	0.4557	-7470A	PPB	04 Aug 2020 11:25:55	U	SMPL
L3525-02DUP - 1	TILCON-FL-002DUP	1402	0.4541	-7470A	PPB	04 Aug 2020 11:35:24	U	SMPL
L3525-02MS - 1	TILCON-FL-002MS	15242	4.3417	-7470A	PPB	04 Aug 2020 11:37:30	U	SMPL
CCV46 - 1	CCV46	18040	5.1277	-7470A	PPB	04 Aug 2020 11:39:36	U	SMPL
CCB46 - 1	CCB46	-82	0.0372	-7470A	PPB	04 Aug 2020 11:41:45	U	SMPL
L3525-02MSD - 1	TILCON-FL-002MSD	15549	4.4280	-7470A	PPB	04 Aug 2020 11:43:51	U	SMPL
PB130709BL - 1	PBW	-142	0.0203	-7470A	PPB	04 Aug 2020 11:45:58	U	SMPL
PB130709BS - 1	LCSW	13021	3.7178	-7470A	PPB	04 Aug 2020 11:48:06	U	SMPL
L3536-01 - 1	JTY1-GW-01	3025	0.9100	-7470A	PPB	04 Aug 2020 11:54:49	U	SMPL
L3536-01DUP - 1	JTY1-GW-01DUP	2667	0.8094	-7470A	PPB	04 Aug 2020 11:56:55	U	SMPL
L3536-01MS - 1	JTY1-GW-01MS	19186	5.4496	-7470A	PPB	04 Aug 2020 11:59:04	U	SMPL
L3536-01MSD - 1	JTY1-GW-01MSD	18633	5.2942	-7470A	PPB	04 Aug 2020 12:01:12	U	SMPL
L3537-04 - 1	JTY1-WC-L-01	675	0.2498	-7470A	PPB	04 Aug 2020 12:03:21	U	SMPL
L3539-01 - 1	I-OC-5	-9	0.0577	-7470A	PPB	04 Aug 2020 12:05:30	U	SMPL
L3539-05 - 1	I-R-5	39	0.0712	-7470A	PPB	04 Aug 2020 12:07:37	U	SMPL
CCV47 - 1	CCV47	18994	5.3957	-7470A	PPB	04 Aug 2020 12:09:43	U	SMPL
CCB47 - 1	CCB47	-57	0.0442	-7470A	PPB	04 Aug 2020 12:11:50	U	SMPL
L3521-04 - 1	COMP-AB	26	0.0675	-7470A	PPB	04 Aug 2020 12:13:56	U	SMPL
L3539-06 - 1	E-R-5	0	0.0602	-7470A	PPB	04 Aug 2020 12:16:02	U	SMPL
PB130707TB - 1	PB130674TB	23	0.0667	-7470A	PPB	04 Aug 2020 12:18:08	U	SMPL
PB130708TB - 1	PB130672TB	20	0.0659	-7470A	PPB	04 Aug 2020 12:20:15	U	SMPL
L3537-01LX5 - 1		16	0.0647	-7470A	PPB	04 Aug 2020 12:22:21	U	SMPL
L3537-01A - 1		13285	3.7920	-7470A	PPB	04 Aug 2020 12:24:29	U	SMPL
L3525-02LX5 - 1		211	0.1195	-7470A	PPB	04 Aug 2020 12:26:36	U	SMPL
L3525-02A - 1		15584	4.4378	-7470A	PPB	04 Aug 2020 12:28:44	U	SMPL
L3536-01LX5 - 1		587	0.2251	-7470A	PPB	04 Aug 2020 12:30:52	U	SMPL
L3536-01A - 1		17745	5.0448	-7470A	PPB	04 Aug 2020 12:33:00	U	SMPL
CCV48 - 1	CCV48	18592	5.2827	-7470A	PPB	04 Aug 2020 12:35:07	U	SMPL
CCB48 - 1	CCB48	-99	0.0324	-7470A	PPB	04 Aug 2020 12:37:15	U	SMPL