



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110364

Slope : 97.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/04/2020	16:20
2	CAL2	1	Water	NA	NA	21.3	7.00	08/04/2020	16:21
3	CAL3	1	Water	NA	NA	21.3	10.03	08/04/2020	16:22
4	ICV	1	Water	NA	NA	21.4	6.98	08/04/2020	16:25
5	CCV1	1	Water	NA	NA	21.4	2.01	08/04/2020	16:26
6	L3545-01	1	Solid	20.03	20	24.9	10.07	08/04/2020	16:28
7	L3553-04	1	Solid	20.02	20	25.4	8.36	08/04/2020	16:30
8	L3553-04DUP	1	Solid	20.03	20	25.5	8.37	08/04/2020	16:31
9	CCV2	1	Water	NA	NA	21.5	12.02	08/04/2020	16:33

WORKLIST(Hardcopy Internal Chain)

VB110364

WorkList Name : corrosivity I3557 WorkList ID : 141136 Department : Wet-Chemistry Date : 08-04-2020 11:06:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/31/2020	Solid	L3545-01	Corrosivity	Cool 4 deg C	DIST01	I43	WC-1	07/31/2020	9045D
08/03/2020	Solid	L3553-04	Corrosivity	Cool 4 deg C	PSEG01	J12	JB-1A	08/03/2020	9045D

Date/Time 08/04/2020 16:10 (nm)
 Received by: JA Wet
 Relinquished by: JA Wet (nm)

Date/Time 08/04/2020 17:00
 Received by: JA Wet
 Relinquished by: JA Wet (nm)