


LB110368

7470A

INSTRUMENT 1D: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	-36	0.000	-36				
0.2	0.200	0.184	-0.016	514	0.0 %	514				
2.5	2.500	2.423	-0.077	7881	0.0 %	7881				
5.0	5.000	5.132	0.132	16797	0.0 %	16797				
7.5	7.500	7.469	-0.031	24490	0.0 %	24490				
10.0	10.000	9.974	-0.026	32734	0.0 %	32734				

LB110368 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Aug 2020 11:26:50

Method: 7470A		Operator: Admin		Date of Analysis: 05 Aug 2020 11:20:00					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S'0	-36	-	0.0000	7470A	PPB	05 Aug 2020 11:37:16	S	Std
0.2 - 1	S'0.2	514	-	0.2000	7470A	PPB	05 Aug 2020 11:39:24	S	Std
2.5 - 1	S'2.5	7881	-	2.5000	7470A	PPB	05 Aug 2020 11:41:32	S	Std
5.0 - 1	S'5	16797	-	5.0000	7470A	PPB	05 Aug 2020 11:43:39	S	Std
7.5 - 1	S'7.5	24490	-	7.5000	7470A	PPB	05 Aug 2020 11:45:46	S	Std
10.0 - 1	S'10	32734	-	10.0000	7470A	PPB	05 Aug 2020 11:47:53	S	Std
ICV56 - 1	ICV56	12137	3.7160	-7470A	PPB	PPB	05 Aug 2020 11:50:27	U	SMPL
ICB56 - 1	ICB56	-147	-0.0164	-7470A	PPB	PPB	05 Aug 2020 11:54:54	U	SMPL
CCV52 - 1	CCV52	16720	5.1085	-7470A	PPB	PPB	05 Aug 2020 11:57:00	U	SMPL
CCB52 - 1	CCB52	-106	-0.0039	-7470A	PPB	PPB	05 Aug 2020 12:02:23	U	SMPL
CRA - 1	CRA	481	0.1744	-7470A	PPB	PPB	05 Aug 2020 12:04:29	U	SMPL
HighStd - 1	HighStd	31762	9.6788	-7470A	PPB	PPB	05 Aug 2020 12:06:35	U	SMPL
ChkStd - 1	ChkStd	22773	6.9476	-7470A	PPB	PPB	05 Aug 2020 12:08:41	U	SMPL
PB130743BL - 1	PBW	-104	-0.0033	-7470A	PPB	PPB	05 Aug 2020 12:13:25	U	SMPL
PB130743BS - 1	LCSW	11923	3.6510	-7470A	PPB	PPB	05 Aug 2020 12:15:31	U	SMPL
L3365-06 - 1	52-09-57	-280	-0.0568	-7470A	PPB	PPB	05 Aug 2020 12:17:38	U	SMPL
L3547-02 - 1	CRT-012	-119	-0.0079	-7470A	PPB	PPB	05 Aug 2020 12:19:46	U	SMPL
L3547-02DUP - 1	CRT-012DUP	-87	0.0019	-7470A	PPB	PPB	05 Aug 2020 12:21:54	U	SMPL
L3547-02MS - 1	CRT-012MS	12434	3.8062	-7470A	PPB	PPB	05 Aug 2020 12:24:02	U	SMPL
L3547-02MSD - 1	CRT-012MSD	12670	3.8779	-7470A	PPB	PPB	05 Aug 2020 12:26:08	U	SMPL
CCV53 - 1	CCV53	17866	5.4567	-7470A	PPB	PPB	05 Aug 2020 12:28:14	U	SMPL
CCB53 - 1	CCB53	-103	-0.0030	-7470A	PPB	PPB	05 Aug 2020 12:32:48	U	SMPL
L3548-02 - 1	CRT-025	-89	0.0013	-7470A	PPB	PPB	05 Aug 2020 12:34:54	U	SMPL
L3548-05 - 1	CRT-017	-73	0.0061	-7470A	PPB	PPB	05 Aug 2020 12:37:00	U	SMPL
L3548-08 - 1	CRT-004	-48	0.0137	-7470A	PPB	PPB	05 Aug 2020 12:39:06	U	SMPL
L3548-11 - 1	CRT-019	-43	0.0152	-7470A	PPB	PPB	05 Aug 2020 12:41:12	U	SMPL
L3551-03 - 1	SP-1-DARK-SOIL	-32	0.0186	-7470A	PPB	PPB	05 Aug 2020 12:43:18	U	SMPL
PB130743TB - 1	PB130719TB	-17	0.0231	-7470A	PPB	PPB	05 Aug 2020 12:45:25	U	SMPL
L3547-02LX5 - 1		-51	0.0128	-7470A	PPB	PPB	05 Aug 2020 12:47:32	U	SMPL
L3547-02A - 1		12464	3.8153	-7470A	PPB	PPB	05 Aug 2020 12:49:40	U	SMPL
CCV54 - 1	CCV54	17261	5.2729	-7470A	PPB	PPB	05 Aug 2020 12:51:49	U	SMPL
CCB54 - 1	CCB54	52	0.0441	-7470A	PPB	PPB	05 Aug 2020 12:56:15	U	SMPL