



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110371

Slope : 98.2

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/05/2020	11:40
2	CAL2	1	Water	NA	NA	21.3	7.01	08/05/2020	11:41
3	CAL3	1	Water	NA	NA	21.3	10.03	08/05/2020	11:42
4	ICV	1	Water	NA	NA	21.3	7.00	08/05/2020	11:45
5	CCV1	1	Water	NA	NA	21.4	2.01	08/05/2020	11:48
6	L3571-01	1	Water	NA	NA	25.6	5.06	08/05/2020	11:50
7	L3571-01DUP	1	Water	NA	NA	25.7	5.07	08/05/2020	11:51
8	L3571-02	1	Water	NA	NA	25.5	4.91	08/05/2020	12:00
9	L3571-03	1	Water	NA	NA	25.6	4.92	08/05/2020	12:01
10	CCV2	1	Water	NA	NA	21.4	12.02	08/05/2020	12:06

LB 110371

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3571

Department : Wet-Chemistry

WorkList ID : 141170

Date : 08-05-2020 11:34:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/04/2020	Water	L3571-01-17	pH	Cool 4 deg C	STER13	J12	CELL-A-POOL-WATER	08/04/2020	SM 4500-H B
08/04/2020	Water	L3571-02-17	pH	Cool 4 deg C	STER13	J12	CELL-B-POOL-WATER	08/04/2020	SM 4500-H B
08/04/2020	Water	L3571-03-17	pH	Cool 4 deg C	STER13	J12	CELL-C-POOL-WATER	08/04/2020	SM 4500-H B

Date/Time 08/05/2020 11:35
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 08/05/2020 13:40
 Received by: [Signature]
 Relinquished by: [Signature]