



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110372

Slope : 98.2

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	08/05/2020	11:40
2	CAL2	1	Water	NA	NA	21.3	7.01	08/05/2020	11:41
3	CAL3	1	Water	NA	NA	21.3	10.03	08/05/2020	11:42
4	ICV	1	Water	NA	NA	21.3	7.00	08/05/2020	11:45
5	CCV1	1	Water	NA	NA	21.4	2.01	08/05/2020	11:48
6	L3537-04	1	Water	NA	NA	16.1	8.19	08/05/2020	12:10
7	L3537-04DUP	1	Water	NA	NA	16.2	8.20	08/05/2020	12:11
8	CCV2	1	Water	NA	NA	21.3	12.03	08/05/2020	12:12

LB110372

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH L3537 WorkList ID: 141171 Department: Wet-Chemistry Date: 08-05-2020 11:35:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2020	Water	L3537-04	pH	Cool 4 deg C	PARS15	I53	JTY1-WC-L-01	07/30/2020	9040C

Date/Time 08/05/2020 11:37
 Received by: J. Wet-Chem
 Relinquished by: J. Wet-Chem

Date/Time 08/05/2020 13:45
 Received by: J. Wet-Chem
 Relinquished by: J. Wet-Chem