



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110380

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	08/05/2020	16:30
2	CAL2	1	Water	NA	NA	21.1	7.00	08/05/2020	16:31
3	CAL3	1	Water	NA	NA	21.1	10.03	08/05/2020	16:32
4	ICV	1	Water	NA	NA	21.1	7.01	08/05/2020	16:33
5	CCV1	1	Water	NA	NA	21.2	2.01	08/05/2020	16:35
6	L3502-05	1	Solid	20.03	20	24.1	8.76	08/05/2020	16:37
7	L3502-05DUP	1	Solid	20.03	20	24.2	8.77	08/05/2020	16:38
8	CCV2	1	Water	NA	NA	21.2	12.03	08/05/2020	16:40

VB 110380

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I3502 WorkList ID : 141186 Department : Wet-Chemistry Date : 08-05-2020 13:34:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2020	Solid	L-3502-05	Corrosivity	Cool 4 deg C	PSEG05	J13	OK-03-080520	08/05/2020	9045D

Date/Time 08/05/2020 15:00
 Received by: JD (Wet-Chem)
 Relinquished by: DP (Digestion Lab)

Date/Time 08/05/2020 17:00
 Received by: DP (Digestion Lab)
 Relinquished by: DP (Wet-Chem)