



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110382

Slope : 98.3

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/06/2020	10:20
2	CAL2	1	Water	NA	NA	21.2	7.00	08/06/2020	10:21
3	CAL3	1	Water	NA	NA	21.2	10.03	08/06/2020	10:22
4	ICV	1	Water	NA	NA	21.1	6.98	08/06/2020	10:24
5	CCV1	1	Water	NA	NA	21.2	2.01	08/06/2020	10:25
6	L3563-03	1	Water	NA	NA	19.6	6.40	08/06/2020	10:26
7	L3563-03DUP	1	Water	NA	NA	19.7	6.41	08/06/2020	10:27
8	L3563-04	1	Water	NA	NA	20.1	6.53	08/06/2020	10:30
9	L3563-07	1	Water	NA	NA	19.5	5.00	08/06/2020	10:31
10	L3563-08	1	Water	NA	NA	20.1	5.65	08/06/2020	10:32
11	CCV2	1	Water	NA	NA	21.2	12.02	08/06/2020	10:35

LB 110382

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3563

Department : Wet-Chemistry

WorkList ID : 141199

Date : 08-06-2020 10:09:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2020	Water	L3563-03	pH	Cool 4 deg C	PSEG01	J13	NWM-1509	08/05/2020	9040C
08/05/2020	Water	L3563-04	pH	Cool 4 deg C	PSEG01	J13	NWM-1510	08/05/2020	9040C
08/05/2020	Water	L3563-07	pH	Cool 4 deg C	PSEG01	J13	NMB-15:06	08/05/2020	9040C
08/05/2020	Water	L3563-08	pH	Cool 4 deg C	PSEG01	J13	NWB-15:08	08/05/2020	9040C

Date/Time 08/06/2020 10:15
 Received by: J. Wet-Chem
 Relinquished by: CP (sm)

Date/Time 08/06/2020 11:35
 Received by: CP (sm)
 Relinquished by: J.D. Wet-Chem