



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110383

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/06/2020	10:20
2	CAL2	1	Water	NA	NA	21.2	7.00	08/06/2020	10:21
3	CAL3	1	Water	NA	NA	21.2	10.03	08/06/2020	10:22
4	ICV	1	Water	NA	NA	21.1	6.98	08/06/2020	10:24
5	CCV1	1	Water	NA	NA	21.2	2.01	08/06/2020	10:25
6	L3563-01	1	Solid	20.03	20	22.3	6.52	08/06/2020	10:40
7	L3566-02	1	Solid	20.04	20	21.9	5.54	08/06/2020	10:41
8	L3566-03	1	Solid	20.03	20	22.2	8.96	08/06/2020	10:43
9	L3566-04	1	Solid	20.04	20	21.9	8.96	08/06/2020	10:45
10	L3566-04DUP	1	Solid	20.03	20	21.8	8.97	08/06/2020	10:46
11	CCV2	1	Water	NA	NA	21.2	12.02	08/06/2020	10:49

20110383

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L3563 WorkList ID : 141200 Department : Wet-Chemistry Date : 08-06-2020 10:10:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2020	Solid	L3563-01	pH	Cool 4 deg C	PSEG01	J13	NB-30860	08/05/2020	9045D
08/05/2020	Solid	L3566-02	pH	Cool 4 deg C	PSEG01	J21	31161	08/05/2020	9045D
08/05/2020	Solid	L3566-03	pH	Cool 4 deg C	PSEG01	J21	31162	08/05/2020	9045D
08/05/2020	Solid	L3566-04	pH	Cool 4 deg C	PSEG01	J21	31163	08/05/2020	9045D

Date/Time 08/06/2020 10:15
Received by: JP (wet) (Chen)
Relinquished by: JP (sm)

Date/Time 08/06/2020 11:35
Received by: JP (sm)
Relinquished by: JP (wet) (Chen)