


LB110384

ISM02.4

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.6529e-004

C= -3.7500e-002

Rho= 0.9994664

Accept=Accepted

*slope
y=1015xcept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9.11
0.00	0.000	-0.030	-0.030	27	0.000	27					11
0.05	0.050					294					22
0.20	0.200	0.243	0.043	1059	0.0 %	1059					3
2.50	2.500	2.431	-0.069	9306	0.0 %	9306					30
5.00	5.000	4.979	-0.021	18910	0.0 %	18910					3
7.50	7.500	7.733	0.233	29292	0.0 %	29292					2
10.0	10.000	9.843	-0.157	37243	0.0 %	37243					2

LB110384
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	27	-		0 ISM02.4	PPB	8/6/2020 11:30	Std
	0.2 S0.2	1059	-		0.2 ISM02.4	PPB	8/6/2020 11:33	Std
	2.5 S2.5	9306	-		2.5 ISM02.4	PPB	8/6/2020 11:35	Std
	5 S5.0	18910	-		5 ISM02.4	PPB	8/6/2020 11:37	Std
	7.5 S7.5	29292	-		7.5 ISM02.4	PPB	8/6/2020 11:40	Std
	10 S10.0	37243	-		10 ISM02.4	PPB	8/6/2020 11:50	Std
ICV	ICV	14692	3.8602	-	ISM02.4	PPB	8/6/2020 11:53	SMPL
ICB	ICB	-10	-0.0402	-	ISM02.4	PPB	8/6/2020 11:58	SMPL
CCV058	CCV058	19424	5.1156	-	ISM02.4	PPB	8/6/2020 12:00	SMPL
CCB058	CCB058	-143	-0.0754	-	ISM02.4	PPB	8/6/2020 12:02	SMPL
PB130756BL	PBW756	-88	-0.0608	-	ISM02.4	PPB	8/6/2020 12:05	SMPL
L3514-02	PK-10S	-68	-0.0555	-	ISM02.4	PPB	8/6/2020 12:07	SMPL
L3514-03	PK-10D	519	0.1002	-	ISM02.4	PPB	8/6/2020 12:09	SMPL
L3514-06	PK-10I	194	0.014	-	ISM02.4	PPB	8/6/2020 12:11	SMPL
L3514-07	RW-12I	150	0.0023	-	ISM02.4	PPB	8/6/2020 12:14	SMPL
L3514-08	RW-12D	11	-0.0346	-	ISM02.4	PPB	8/6/2020 12:16	SMPL
L3514-09	PK-10S	-42	-0.0486	-	ISM02.4	PPB	8/6/2020 12:18	SMPL
L3514-10	PK-10D	472	0.0877	-	ISM02.4	PPB	8/6/2020 12:20	SMPL
L3514-11	L3514-10D	483	0.0906	-	ISM02.4	PPB	8/6/2020 12:23	SMPL
L3514-12	L3514-10S	4473	1.1492	-	ISM02.4	PPB	8/6/2020 12:25	SMPL
L3514-13	PK-10I	-26	-0.0444	-	ISM02.4	PPB	8/6/2020 12:27	SMPL
L3514-14	RW-12I	113	-0.0075	-	ISM02.4	PPB	8/6/2020 12:29	SMPL
L3514-15	RW-12D	11	-0.0346	-	ISM02.4	PPB	8/6/2020 12:32	SMPL
L3514-16	L3514-03D	502	0.0957	-	ISM02.4	PPB	8/6/2020 12:34	SMPL
L3514-17	L3514-03S	4441	1.1407	-	ISM02.4	PPB	8/6/2020 12:36	SMPL
PB130757BL	PBW757	-80	-0.0587	-	ISM02.4	PPB	8/6/2020 12:39	SMPL
L3528-01	METKY9	-8	-0.0396	-	ISM02.4	PPB	8/6/2020 12:41	SMPL
L3528-02	METYL1	524	0.1015	-	ISM02.4	PPB	8/6/2020 12:43	SMPL
L3528-03	METYL2	257	0.0307	-	ISM02.4	PPB	8/6/2020 12:45	SMPL
L3528-04	METYL5	472	0.0877	-	ISM02.4	PPB	8/6/2020 12:48	SMPL
CCV059	CCV059	19580	5.157	-	ISM02.4	PPB	8/6/2020 12:50	SMPL
CCB059	CCB059	83	-0.0155	-	ISM02.4	PPB	8/6/2020 12:57	SMPL
L3528-05	METYL5D	471	0.0875	-	ISM02.4	PPB	8/6/2020 12:59	SMPL
L3528-06	METYL5S	4823	1.242	-	ISM02.4	PPB	8/6/2020 13:02	SMPL
L3528-07	METYL7	-163	-0.0807	-	ISM02.4	PPB	8/6/2020 13:04	SMPL
L3528-08	METYL9	347	0.0546	-	ISM02.4	PPB	8/6/2020 13:06	SMPL
L3528-09	METYM1	1002	0.2283	-	ISM02.4	PPB	8/6/2020 13:08	SMPL
PB130758BL	PBW758	-183	-0.086	-	ISM02.4	PPB	8/6/2020 13:11	SMPL
L3544-01	MJLW61	5	-0.0362	-	ISM02.4	PPB	8/6/2020 13:13	SMPL
L3544-02	MJLW61D	9	-0.0351	-	ISM02.4	PPB	8/6/2020 13:15	SMPL
L3544-03	MJLW61S	4296	1.1022	-	ISM02.4	PPB	8/6/2020 13:17	SMPL
L3544-04	MJLW63	0	-0.0375	-	ISM02.4	PPB	8/6/2020 13:20	SMPL
L3544-05	MJLW65	281	0.037	-	ISM02.4	PPB	8/6/2020 13:22	SMPL
L3544-06	MJLW67	177	0.0095	-	ISM02.4	PPB	8/6/2020 13:24	SMPL
L3544-07	MJLW69	124	-0.0046	-	ISM02.4	PPB	8/6/2020 13:27	SMPL
L3544-08	MJLW71	209	0.0179	-	ISM02.4	PPB	8/6/2020 13:29	SMPL

LB110384
INSTRUMENT ID : CV1

L3544-09	MJLW87	1	-0.0372 -	ISM02.4	PPB	8/6/2020 13:31	SMPL
L3544-10	MJLW88	-16	-0.0417 -	ISM02.4	PPB	8/6/2020 13:33	SMPL
L3544-11	MJLW89	0	-0.0375 -	ISM02.4	PPB	8/6/2020 13:36	SMPL
L3544-12	MJLW91	12	-0.0343 -	ISM02.4	PPB	8/6/2020 13:38	SMPL
PB130759BL	PBW759	16	-0.0333 -	ISM02.4	PPB	8/6/2020 13:40	SMPL
L3546-01	MJLW62	25	-0.0309 -	ISM02.4	PPB	8/6/2020 13:42	SMPL
CCV060	CCV060	19484	5.1315 -	ISM02.4	PPB	8/6/2020 13:45	SMPL
CCB060	CCB060	194	0.014 -	ISM02.4	PPB	8/6/2020 13:47	SMPL
L3546-02	MJLW62D	-5	-0.0388 -	ISM02.4	PPB	8/6/2020 13:49	SMPL
L3546-03	MJLW62S	4342	1.1144 -	ISM02.4	PPB	8/6/2020 13:51	SMPL
L3546-04	MJLW64	21	-0.0319 -	ISM02.4	PPB	8/6/2020 13:54	SMPL
L3546-05	MJLW66	68	-0.0195 -	ISM02.4	PPB	8/6/2020 13:56	SMPL
L3546-06	MJLW68	91	-0.0134 -	ISM02.4	PPB	8/6/2020 13:58	SMPL
L3546-07	MJLW70	44	-0.0258 -	ISM02.4	PPB	8/6/2020 14:01	SMPL
L3546-08	MJLW72	14	-0.0338 -	ISM02.4	PPB	8/6/2020 14:03	SMPL
CCV061	CCV061	19436	5.1188 -	ISM02.4	PPB	8/6/2020 14:05	SMPL
CCB061	CCB061	92	-0.0131 -	ISM02.4	PPB	8/6/2020 14:07	SMPL