



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110386

Slope : 98.3

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/06/2020	10:20
2	CAL2	1	Water	NA	NA	21.2	7.00	08/06/2020	10:21
3	CAL3	1	Water	NA	NA	21.2	10.03	08/06/2020	10:22
4	ICV	1	Water	NA	NA	21.1	6.98	08/06/2020	10:24
5	CCV1	1	Water	NA	NA	21.2	2.01	08/06/2020	10:25
6	L3584-01	1	Water	NA	NA	13.5	6.74	08/06/2020	11:00
7	L3587-01	1	Water	NA	NA	25.5	5.87	08/06/2020	11:01
8	L3587-01DUP	1	Water	NA	NA	25.6	5.88	08/06/2020	11:02
9	CCV2	1	Water	NA	NA	21.2	12.02	08/06/2020	11:05

WORKLIST(Hardcopy Internal Chain)

AB 110386

WorkList Name : PH L3587

WorkList ID : 141213

Department : Wet-Chemistry

Date : 08-06-2020 11:17:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/05/2020	Water	L3584-01	✓ pH	Cool 4 deg C	APT01	J21	WGGM-20-WW-TSF-026	08/05/2020	SM 4500-H B
08/03/2020	Water	L3587-01	✗ pH	Cool 4 deg C	STER06	J21	POOL-WATER	08/03/2020	SM 4500-H B

Date/Time 08/06/2020 11:20
 Received by: Jo (wet-chem)
 Relinquished by: AMLwet-chem

Date/Time 08/06/2020 13:40
 Received by: AMLwet-chem
 Relinquished by: JH (wet-chem)