



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB110387

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP86087
Turbidity Calibration std, 20NTU	WP86083
Turbidity Calibration std, 40NTU	WP86082
Turbidity Calibration std, 1NTU	WP86086
Turbidity Calibration std, 5NTU	WP86085
Turbidity Calibration - CCV std, 10 NTU	WP86084
10 NTU Standard 500 ml	W2489

Intercept: 0.2824

Slope: 0.9631

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.233	08/06/2020	11:20
2	CAL2	1	1	1.200	08/06/2020	11:23
3	CAL3	5	1	5.140	08/06/2020	11:26
4	CAL4	10	1	9.820	08/06/2020	11:30
5	CAL5	20	1	19.800	08/06/2020	11:32
6	CAL6	40	1	38.700	08/06/2020	11:35



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB110387

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.298	08/06/2020	11:38
2	ICB		1	0.252	-0.032	08/06/2020	11:40
3	CCV1	10	1	9.920	10.007	08/06/2020	11:42
4	CCB1		1	0.245	-0.039	08/06/2020	11:45
5	LB110387BL		1	0.267	-0.016	08/06/2020	11:48
6	L3585-01		1	15.200	15.489	08/06/2020	11:50
7	L3585-01DUP		1	14.700	14.970	08/06/2020	11:52
8	CCV2	10	1	9.900	9.986	08/06/2020	11:55
9	CCB2		1	0.272	-0.011	08/06/2020	11:57

LB110387

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-80620 WorkList ID : 141212 Department : Wet-Chemistry Date : 08-06-2020 11:04:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2020	Water	L3585-01	Turbidity	Cool 4 deg C	LIRO01	J21	UDS-6-16A	08/05/2020	SM2130 B

Date/Time 8/06/2020 11:16

Received by: AML wet-chem

Relinquished by: AP (wet-chem)

Date/Time 8/6/2020 12:05

Received by: AP (wet-chem)

Relinquished by: AML (wet-chem)