



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB110396

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/06/2020	16:30
2	CAL2	1	Water	NA	NA	21.2	7.00	08/06/2020	16:31
3	CAL3	1	Water	NA	NA	21.2	10.03	08/06/2020	16:33
4	ICV	1	Water	NA	NA	21.2	7.00	08/06/2020	16:35
5	CCV1	1	Water	NA	NA	21.3	2.01	08/06/2020	16:36
6	L3508-04	1	Solid	20.03	20	24.1	8.41	08/06/2020	16:39
7	L3508-04DUP	1	Solid	20.04	20	24.2	8.42	08/06/2020	16:40
8	L3594-02	1	Solid	20.01	20	23.3	5.36	08/06/2020	16:41
9	L3594-06	1	Solid	20.03	20	22.9	5.56	08/06/2020	16:43
10	L3594-10	1	Solid	20.04	20	22.5	7.90	08/06/2020	16:45
11	CCV2	1	Water	NA	NA	21.2	12.02	08/06/2020	16:46

✓ 110396

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity I3594 WorkList ID : 141219 Department : Wet-Chemistry Date : 08-06-2020 12:53:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/06/2020	Solid	L3508-04	Corrosivity	Cool 4 deg C	PSEG05	J21	TR-04-080620	08/06/2020	9045D
08/05/2020	Solid	L3594-02	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-1	08/05/2020	9045D
08/05/2020	Solid	L3594-06	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-2	08/05/2020	9045D
08/05/2020	Solid	L3594-10	Corrosivity	Cool 4 deg C	PSEG01	J21	TP-3	08/05/2020	9045D

Date/Time 08/06/2020 15:40
 Received by: Jg. Ault-6rem
 Relinquished by: CP (SMI)

Date/Time 08/06/2020 17:35
 Received by: CP (SMI)
 Relinquished by: Jg. Ault-6rem