



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 8/7/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:27
In Date: 08/06/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 08/07/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB110397

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L3598-01	GAY59	1	1.15	9.91	8.02	78.4	
L3598-02	GAY60	2	1.13	9.66	7.85	78.8	
L3598-03	GAY61	3	1.14	9.54	8.23	84.4	
L3598-04	GAY62	4	1.15	9.95	8.29	81.1	
L3598-05	GAY63	5	1.14	9.52	7.87	80.3	
L3598-06	GAY64	6	1.16	9.79	8.06	80.0	
L3598-07	GAY64MS	7	1.16	9.79	8.06	80.0	
L3598-08	GAY64MSD	8	1.16	9.79	8.06	80.0	
L3598-09	GAY65	9	1.19	9.75	7.85	77.8	
L3598-10	GAY66	10	1.16	9.96	7.91	76.7	
L3598-11	GAY67	11	1.11	9.97	8.66	85.2	
L3598-12	GAY68	12	1.17	9.73	8.18	81.9	
L3598-13	GAY69	13	1.12	9.91	8.55	84.5	
L3598-14	GAY70	14	1.18	9.98	8.39	81.9	
L3598-15	GAY71	15	1.14	9.68	7.62	75.9	
L3598-16	GAY72	16	1.17	9.95	7.86	76.2	
L3598-17	GAY73	17	1.17	9.88	8.04	78.9	
L3598-18	GAY74	18	1.14	9.94	7.8	75.7	
L3598-19	GAY75	19	1.17	9.73	7.58	74.9	
L3598-20	GAY76	20	1.18	9.75	8.13	81.1	
L3598-21	GAY77	21	1.18	9.96	8.15	79.4	
L3598-22	GAY78	22	1.15	9.96	9.00	89.1	
L3598-23	GAY81	23	1.00	2.00	2.00	100.0	P.E. SAMPLE
L3598-24	GAY82	24	1.00	2.00	2.00	100.0	P.E. SAMPLE
L3598-25	VHBLK01	25	1.00	2.00	2.00	100.0	VHBLK

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$