

LB110398


ISM02.4

INSTRUMENT ID: CV1

Linear

μ Abs.:

37930



A= 0.0000e+000

B= 2.6374e-004

C= -6.5812e-002

Rho= 0.9999324

Accept=Accepted

*slope
y-intercept*

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.051	-0.051	57	0.000	57					
0.05	0.050					294					
0.20	0.200	0.217	0.017	1071	0.0 %	1071					
2.50	2.500	2.506	0.006	9753	0.0 %	9753					
5.00	5.000	5.033	0.033	19333	0.0 %	19333					
7.50	7.500	7.557	0.057	28901	0.0 %	28901					
10.0	10.000	9.938	-0.062	37930	0.0 %	37930					

*0/0 D
119011*

LB110398

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	57	-	0	ISM02.4	PPB	8/6/2020 16:07	Std
	0.2 S0.2	1071	-	0.2	ISM02.4	PPB	8/6/2020 16:12	Std
	2.5 S2.5	9753	-	2.5	ISM02.4	PPB	8/6/2020 16:14	Std
	5 S5.0	19333	-	5	ISM02.4	PPB	8/6/2020 16:16	Std
	7.5 S7.5	28901	-	7.5	ISM02.4	PPB	8/6/2020 16:21	Std
	10 S10.0	37930	-	10	ISM02.4	PPB	8/6/2020 16:23	Std
ICV	ICV	14758	3.8265 -		ISM02.4	PPB	8/6/2020 16:26	SMPL
ICB	ICB	-31	-0.074 -		ISM02.4	PPB	8/6/2020 16:28	SMPL
CCV062	CCV062	19606	5.1051 -		ISM02.4	PPB	8/6/2020 16:31	SMPL
CCB062	CCB062	239	-0.0028 -		ISM02.4	PPB	8/6/2020 16:33	SMPL
PB130793BL	PBW793	288	0.0101 -		ISM02.4	PPB	8/6/2020 16:38	SMPL
PB130793TB	LEB793	409	0.0421 -		ISM02.4	PPB	8/6/2020 16:44	SMPL
L3567-01	MBG0A0	1077	0.2182 -		ISM02.4	PPB	8/6/2020 16:51	SMPL
L3567-02	MBG0A0D	1023	0.204 -		ISM02.4	PPB	8/6/2020 16:55	SMPL
L3567-03	MBG0A0S	5280	1.3267 -		ISM02.4	PPB	8/6/2020 16:58	SMPL
CCV063	CCV063	20018	5.2138 -		ISM02.4	PPB	8/6/2020 17:00	SMPL
CCB063	CCB063	6	-0.0642 -		ISM02.4	PPB	8/6/2020 17:02	SMPL