



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB110407

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	08/07/2020	09:25
2	L3563-05		1	8	100.00	>212.0	>212.0	08/07/2020	09:50
3	L3563-05DUP		1	8	100.00	>212.0	>212.0	08/07/2020	10:20
4	L3563-09		1	9	100.00	>212.0	>212.0	08/07/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP8720 WorkList ID : 141249 Department : Wet-Chemistry Date : 08-07-2020 08:59:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2020	Water	L3563-05	Flash Point	Cool 4 deg C	PSEG01	J13	NWM-COMP	08/05/2020	1010A
08/10/2020	Water	L3563-09	Flash Point	Cool 4 deg C	PSEG01	J13	NWB-COMP	08/05/2020	1010A

Date/Time 08/07/2020 09:05
 Received by: KS (weldan)
 Relinquished by: KS (weldan)

Date/Time 08/07/2020 10:50
 Received by: KS (weldan)
 Relinquished by: KS (weldan)

LB110407