


LB110412

7471B INSTRUMENT ID: CV2

μ Abs.:

52725



Linear

A= 0.0000e+000

B= 1.9088e-004

C= 2.1802e-002

Rho= 0.9998673

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.062	0.062	209	0.000	209				
0.2	0.200	0.224	0.024	1058	0.0 %	1058				
2.5	2.500	2.453	-0.047	12735	0.0 %	12735				
5.0	5.000	4.935	-0.065	25739	0.0 %	25739				
7.5	7.500	7.441	-0.059	38870	0.0 %	38870				
10.0	10.000	10.086	0.086	52725	0.0 %	52725				

LB110412 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 07 Aug 2020 12:11:19					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	209	-	0.0000	7471B	PPB	07 Aug 2020 12:12:59	S	Std
0.2 - 1	50.2	1058	-	0.2000	7471B	PPB	07 Aug 2020 12:15:16	S	Std
2.5 - 1	52.5	12735	-	2.5000	7471B	PPB	07 Aug 2020 12:17:33	S	Std
5.0 - 1	55	25739	-	5.0000	7471B	PPB	07 Aug 2020 12:19:50	S	Std
7.5 - 1	57.5	38870	-	7.5000	7471B	PPB	07 Aug 2020 12:22:07	S	Std
10.0 - 1	50	52725	-	10.0000	7471B	PPB	07 Aug 2020 12:24:25	S	Std
ICV59 - 1	ICV59	20385	3.9129	-	7471B	PPB	07 Aug 2020 12:27:25	U	SMPL
ICB59 - 1	ICB59	-103	0.0021	-	7471B	PPB	07 Aug 2020 12:29:41	U	SMPL
CCV69 - 1	CCV69	26285	5.0390	-	7471B	PPB	07 Aug 2020 12:32:00	U	SMPL
CCB69 - 1	CCB69	-111	0.0006	-	7471B	PPB	07 Aug 2020 12:34:16	U	SMPL
CRA - 1	CRA	1011	0.2148	-	7471B	PPB	07 Aug 2020 12:36:34	U	SMPL
HighStd - 1	HighStd	52196	9.9849	-	7471B	PPB	07 Aug 2020 12:38:49	U	SMPL
ChkStd - 1	ChkStd	36115	6.9154	-	7471B	PPB	07 Aug 2020 12:41:05	U	SMPL
PB130796BL - 1	PBS	-350	-0.0450	-	7471B	PPB	07 Aug 2020 12:43:23	U	SMPL
PB130796BS - 1	LCSS	19005	3.6494	-	7471B	PPB	07 Aug 2020 12:45:42	U	SMPL
L3502-01 - 1	OK-03-080520	9444	1.8245	-	7471B	PPB	07 Aug 2020 12:47:58	U	SMPL
L3508-01 - 1	TR-04-080620	17109	3.2875	-	7471B	PPB	07 Aug 2020 12:50:16	U	SMPL
L3553-02 - 1	JB-1A	7443	1.4425	-	7471B	PPB	07 Aug 2020 12:52:38	U	SMPL
L3563-01 - 1	NB-30860	1532	0.3142	-	7471B	PPB	07 Aug 2020 12:54:56	U	SMPL
L3568-02 - 1	8/5-A-E	3656	0.7197	-	7471B	PPB	07 Aug 2020 12:57:13	U	SMPL
CCV70 - 1	CCV70	25865	4.9589	-	7471B	PPB	07 Aug 2020 12:59:28	U	SMPL
CCB70 - 1	CCB70	12	0.0241	-	7471B	PPB	07 Aug 2020 13:01:43	U	SMPL
L3568-05 - 1	8/5-F-J	2954	0.5857	-	7471B	PPB	07 Aug 2020 13:04:00	U	SMPL
L3573-01 - 1	FK-GF-03-016-A	2881	0.5717	-	7471B	PPB	07 Aug 2020 13:06:15	U	SMPL
L3573-03 - 1	FK-GF-03-016-B	3091	0.6118	-	7471B	PPB	07 Aug 2020 13:08:31	U	SMPL
L3574-01 - 1	FK-GF-02-060-A	3216	0.6357	-	7471B	PPB	07 Aug 2020 13:10:47	U	SMPL
L3574-03 - 1	FK-GF-02-060-B	2695	0.5362	-	7471B	PPB	07 Aug 2020 13:13:02	U	SMPL
L3575-01 - 1	FK-0D015-03-016	1147	0.2407	-	7471B	PPB	07 Aug 2020 13:15:18	U	SMPL
L3576-01 - 1	FK-0D015-03-017	1369	0.2831	-	7471B	PPB	07 Aug 2020 13:17:35	U	SMPL
L3576-01DUP - 1	FK-0D015-03-017DUP	1549	0.3175	-	7471B	PPB	07 Aug 2020 13:19:51	U	SMPL
L3576-01MS - 1	FK-0D015-03-017MS	21336	4.0944	-	7471B	PPB	07 Aug 2020 13:22:08	U	SMPL
L3576-01MSD - 1	FK-0D015-03-017MSD	21442	4.1146	-	7471B	PPB	07 Aug 2020 13:24:24	U	SMPL
CCV71 - 1	CCV71	27101	5.1948	-	7471B	PPB	07 Aug 2020 13:26:41	U	SMPL
CCB71 - 1	CCB71	27	0.0270	-	7471B	PPB	07 Aug 2020 13:28:57	U	SMPL
L3586-02 - 1	8420	2854	0.5666	-	7471B	PPB	07 Aug 2020 13:31:13	U	SMPL
L3594-01 - 1	TP-1	1580	0.3234	-	7471B	PPB	07 Aug 2020 13:33:30	U	SMPL
L3594-05 - 1	TP-2	2182	0.4383	-	7471B	PPB	07 Aug 2020 13:35:45	U	SMPL
L3594-09 - 1	TP-3	2692	0.5356	-	7471B	PPB	07 Aug 2020 13:38:01	U	SMPL
L3597-01 - 1	20200446-CORE-2	11984	2.3093	-	7471B	PPB	07 Aug 2020 13:40:17	U	SMPL
L3576-01LX5 - 1		291	0.0773	-	7471B	PPB	07 Aug 2020 13:42:33	U	SMPL
L3576-01A - 1		21010	4.0322	-	7471B	PPB	07 Aug 2020 13:44:50	U	SMPL
CCV72 - 1	CCV72	27107	5.1959	-	7471B	PPB	07 Aug 2020 13:47:06	U	SMPL
CCB72 - 1	CCB72	41	0.0296	-	7471B	PPB	07 Aug 2020 13:49:24	U	SMPL