



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110419

Slope : 98.1

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	3.99	08/07/2020	13:00
2	CAL2	1	Water	NA	NA	21.1	7.00	08/07/2020	13:01
3	CAL3	1	Water	NA	NA	21.1	10.03	08/07/2020	13:02
4	ICV	1	Water	NA	NA	21.3	7.00	08/07/2020	13:05
5	CCV1	1	Water	NA	NA	21.2	2.02	08/07/2020	13:10
6	L3607-01	1	Water	NA	NA	23.1	10.38	08/07/2020	13:11
7	L3607-01DUP	1	Water	NA	NA	23.2	10.39	08/07/2020	13:12
8	CCV2	1	Water	NA	NA	21.2	12.02	08/07/2020	13:15

28110419

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3607 WorkList ID : 141265 Department : Wet-Chemistry Date : 08-07-2020 12:40:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/07/2020	Water	L3607-01-A	pH	Cool 4 deg C	PSEG01	J21	ES-TOTE-COMP	08/07/2020	9040C

Date/Time 08/07/2020 12:50
Received by: Jn Cwt (hpm)
Relinquished by: Amy(wet-chem)

Date/Time 08/07/2020 14:00
Received by: Amy(wet-chem)
Relinquished by: Jn Cwt (hpm)