

Analysis Method: SM2510 BANALYST: AlexanderParameter: ConductanceSUPERVISOR REVIEW BY: apatelRun Number: LB110425Meter ID: WC PH METER-1Thermometer ID: pH MeterTempProbe

Reagent/Standard	Lot/Log #
conductivity standard	WP86113

Seq	Lab ID	True Value (mg/L)	Dilution	Temperature °C	Reading (umho/cm)	Result (umho/cm)	Anal Date	Anal Time
1	LB110425BL		1	26.1	0.92	0.90	08/07/2020	15:26
2	LB110425BS	1413	1	25.9	1426.00	1401.90	08/07/2020	15:29
3	L3605-01		1	24.3	94.20	95.48	08/07/2020	15:34
4	L3605-01DUP		1	24.2	95.70	97.18	08/07/2020	15:38

Note : Conductivity result is corrected to 25°C

Result = $\frac{\text{(Reading)}}{1 + 0.0191 (\text{Temperature} - 25)}$
--

LB110424-LB110425

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-80720 WorkList ID : 141243 Department : Wet-Chemistry Date : 08-07-2020 08:52:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/10/2020	Water	L3605-01 F	Conductance	Cool 4 deg C	DIST01	J21	DW-QUAL-HOGHOUSE	08/06/2020	SM2510 B
08/08/2020	Water	L3605-01 D	Turbidity	Cool 4 deg C	DIST01	J21	DW-QUAL-HOGHOUSE	08/06/2020	SM2130 B

Date/Time 8/07/2020 14:40
 Received by: AM (wet-chem)
 Relinquished by: JA (wet-chem)

Date/Time 8/07/2020 15:50
 Received by: JA (wet-chem)
 Relinquished by: AM (wet-chem)