



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB110428

Slope : 98.3

pH Meter ID : WC OH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	08/07/2020	16:25
2	CAL2	1	Water	NA	NA	21.2	7.00	08/07/2020	16:26
3	CAL3	1	Water	NA	NA	21.2	10.03	08/07/2020	16:27
4	ICV	1	Water	NA	NA	21.2	6.98	08/07/2020	16:30
5	CCV1	1	Water	NA	NA	21.3	2.01	08/07/2020	16:31
6	L3621-01	1	Water	NA	NA	24.2	5.48	08/07/2020	16:33
7	L3621-01DUP	1	Water	NA	NA	24.3	5.49	08/07/2020	16:34
8	L3621-02	1	Water	NA	NA	24.0	5.15	08/07/2020	16:36
9	L3621-03	1	Water	NA	NA	24.0	4.97	08/07/2020	16:39
10	CCV2	1	Water	NA	NA	21.3	12.02	08/07/2020	16:40

07/10/28

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L3621 WorkList ID : 141280 Department : Wet-Chemistry Date : 08-07-2020 16:13:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/30/2020	Water	L3621-01 - 1	pH	Cool 4 deg C	STER18	J23	JS7500	07/30/2020	SM 4500-H B
07/30/2020	Water	L3621-02 - 1	pH	Cool 4 deg C	STER18	J23	JS9600	07/30/2020	SM 4500-H B
07/30/2020	Water	L3621-03 - 1	pH	Cool 4 deg C	STER18	J23	GAMMA-JARINU	07/30/2020	SM 4500-H B

Date/Time 08/07/2020 16:20
 Received by: JS (wet-chem)
 Relinquished by: JS (wet-chem)

Date/Time 08/07/2020 17:30
 Received by: KS (wet-chem)
 Relinquished by: JS (wet-chem)